

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
 Data File : VN076241.D
 Acq On : 13 Jan 2023 12:14
 Operator : JC\MD
 Sample : PB150223TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150223TB

Quant Time: Jan 16 01:05:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	414602	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	677100	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	582170	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	223807	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	226390	50.192	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.380%		
35) Dibromofluoromethane	8.177	113	204293	48.720	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.440%		
50) Toluene-d8	10.571	98	789100	50.332	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.660%		
62) 4-Bromofluorobenzene	12.853	95	234859	45.229	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 90.460%		
Target Compounds						Qvalue
16) Acetone	4.453	43	50281	32.338	ug/l	99
20) Methylene Chloride	5.294	84	8438	1.406	ug/l	88
37) Ethyl Acetate	7.577	43	39043	8.726	ug/l	98
43) Isopropyl Acetate	8.694	43	205336	27.821	ug/l #	87
95) Naphthalene	15.641	128	10723	3.020	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
Data File : VN076241.D
Acq On : 13 Jan 2023 12:14
Operator : JC\MD
Sample : PB150223TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150223TB

Quant Time: Jan 16 01:05:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

