

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
 Data File : VN076242.D
 Acq On : 13 Jan 2023 12:38
 Operator : JC\MD
 Sample : 01124-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-C

Quant Time: Jan 16 01:06:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	361193	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	583725	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	520568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	228401	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	199777	50.841	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.680%	
35) Dibromofluoromethane	8.171	113	175056	48.426	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.860%	
50) Toluene-d8	10.571	98	688688	50.954	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.900%	
62) 4-Bromofluorobenzene	12.853	95	221243	49.422	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.840%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	51975	38.371	ug/l #	84
37) Ethyl Acetate	7.571	43	33125	8.588	ug/l	99
43) Isopropyl Acetate	8.694	43	234533	36.860	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
Data File : VN076242.D
Acq On : 13 Jan 2023 12:38
Operator : JC\MD
Sample : 01124-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-C

Quant Time: Jan 16 01:06:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

