

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
 Data File : VN076243.D
 Acq On : 13 Jan 2023 13:02
 Operator : JC\MD
 Sample : 01124-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-4

Quant Time: Jan 16 01:06:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	396833	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	642932	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	557766	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	227581	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	218969	50.720	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.440%
35) Dibromofluoromethane	8.177	113	194350	48.812	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.620%
50) Toluene-d8	10.571	98	753322	50.604	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.200%
62) 4-Bromofluorobenzene	12.853	95	228441	46.331	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.660%
Target Compounds						
						Qvalue
16) Acetone	4.448	43	58197	39.105	ug/l	93
37) Ethyl Acetate	7.571	43	33475	7.879	ug/l	99
43) Isopropyl Acetate	8.694	43	280181	39.979	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
Data File : VN076243.D
Acq On : 13 Jan 2023 13:02
Operator : JC\MD
Sample : 01124-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-4

Quant Time: Jan 16 01:06:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

