

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
 Data File : VN076255.D
 Acq On : 13 Jan 2023 17:51
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 16 01:14:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	353221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	560674	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	480720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	186587	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	190513	49.578	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.160%
35) Dibromofluoromethane	8.177	113	167230	48.163	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.320%
50) Toluene-d8	10.571	98	641394	49.406	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.820%
62) 4-Bromofluorobenzene	12.853	95	191696	44.583	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.160%
Target Compounds						
95) Naphthalene	15.641	128	9085	3.033	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
Data File : VN076255.D
Acq On : 13 Jan 2023 17:51
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 16 01:14:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

