

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
 Data File : VN076257.D
 Acq On : 13 Jan 2023 18:39
 Operator : JC\MD
 Sample : 01160-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-2023-1-13

Quant Time: Jan 16 01:15:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	393944	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	621466	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	528086	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	214786	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	203178	47.408	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.820%
35) Dibromofluoromethane	8.177	113	185972	48.321	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.640%
50) Toluene-d8	10.571	98	710281	49.360	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.720%
62) 4-Bromofluorobenzene	12.853	95	214484	45.003	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
Data File : VN076257.D
Acq On : 13 Jan 2023 18:39
Operator : JC\MD
Sample : 01160-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-2023-1-13

Quant Time: Jan 16 01:15:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

