

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
 Data File : VN076258.D
 Acq On : 13 Jan 2023 19:03
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/16/2023
 Supervised By : Mahesh Dadoda 01/16/2023

Quant Time: Jan 16 01:15:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	377960	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	596498	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	531544	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	259903	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	185316	45.069	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.140%		
35) Dibromofluoromethane	8.171	113	172934	46.815	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.620%		
50) Toluene-d8	10.571	98	613105	44.391	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	88.780%		
62) 4-Bromofluorobenzene	12.853	95	214705	46.935	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.860%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.136	85	151472	56.443	ug/l	99
3) Chloromethane	2.377	50	143549	43.017	ug/l	98
4) Vinyl Chloride	2.524	62	184788	46.422	ug/l	96
5) Bromomethane	2.948	94	150336	44.439	ug/l	99
6) Chloroethane	3.118	64	139442	47.304	ug/l	96
7) Trichlorofluoromethane	3.506	101	267093	44.219	ug/l	100
8) Diethyl Ether	3.971	74	97441	44.168	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.383	101	153874	44.613	ug/l	99
10) Methyl Iodide	4.600	142	239531	42.793	ug/l	98
11) Tert butyl alcohol	5.536	59	123727	212.654	ug/l	99
12) 1,1-Dichloroethene	4.353	96	143228	42.757	ug/l	98
13) Acrolein	4.195	56	160599	225.151	ug/l	100
14) Allyl chloride	5.036	41	168700	43.081	ug/l	89
15) Acrylonitrile	5.730	53	350584	220.804	ug/l	100
16) Acetone	4.442	43	275601	194.436	ug/l	97
17) Carbon Disulfide	4.724	76	307080	37.862	ug/l	100
18) Methyl Acetate	5.042	43	198609	47.403	ug/l	99
19) Methyl tert-butyl Ether	5.806	73	531041	48.140	ug/l	99
20) Methylene Chloride	5.289	84	166666	43.876	ug/l	96
21) trans-1,2-Dichloroethene	5.795	96	152430	41.550	ug/l	99
22) Diisopropyl ether	6.683	45	436112	45.845	ug/l	98
23) Vinyl Acetate	6.612	43	1420609	221.873	ug/l	99
24) 1,1-Dichloroethane	6.577	63	279754	44.096	ug/l	98
25) 2-Butanone	7.489	43	433057	218.201	ug/l	95
26) 2,2-Dichloropropane	7.500	77	232571	43.288	ug/l	97
27) cis-1,2-Dichloroethene	7.494	96	192183	43.802	ug/l	98
28) Bromochloromethane	7.818	49	111894	40.240	ug/l	98
29) Tetrahydrofuran	7.847	42	288664	232.094	ug/l	98
30) Chloroform	7.971	83	310534	43.474	ug/l	95
31) Cyclohexane	8.259	56	240462	40.833	ug/l	96
32) 1,1,1-Trichloroethane	8.171	97	279479	44.133	ug/l	98
36) 1,1-Dichloropropene	8.377	75	217439	44.311	ug/l	100
37) Ethyl Acetate	7.565	43	183442	46.540	ug/l	98
38) Carbon Tetrachloride	8.371	117	248787	45.298	ug/l	99
39) Methylcyclohexane	9.606	83	270665	45.720	ug/l	99
40) Benzene	8.612	78	668387	44.146	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
 Data File : VN076258.D
 Acq On : 13 Jan 2023 19:03
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jan 16 01:15:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/16/2023
 Supervised By : Mahesh Dadoda 01/16/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.789	41	98729	48.188	ug/l	98
42) 1,2-Dichloroethane	8.677	62	227652	45.228	ug/l	99
43) Isopropyl Acetate	8.694	43	324655	49.931	ug/l	98
44) Trichloroethene	9.359	130	183287	42.972	ug/l	100
45) 1,2-Dichloropropane	9.624	63	164452	44.636	ug/l	100
46) Dibromomethane	9.712	93	119948	43.894	ug/l	99
47) Bromodichloromethane	9.894	83	242866	45.321	ug/l	98
48) Methyl methacrylate	9.682	41	148148	50.589	ug/l	97
49) 1,4-Dioxane	9.694	88	62141	863.138	ug/l #	96
51) 4-Methyl-2-Pentanone	10.447	43	933892	237.502	ug/l	99
52) Toluene	10.630	92	449091	45.545	ug/l	99
53) t-1,3-Dichloropropene	10.841	75	246973	47.081	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	271557	47.149	ug/l	97
55) 1,1,2-Trichloroethane	11.018	97	177996	45.991	ug/l	97
56) Ethyl methacrylate	10.877	69	253114	49.237	ug/l	97
57) 1,3-Dichloropropane	11.165	76	290018	45.683	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.165	63	409515	250.203	ug/l	100
59) 2-Hexanone	11.200	43	676074	235.763	ug/l	100
60) Dibromochloromethane	11.359	129	199825	46.426	ug/l	99
61) 1,2-Dibromoethane	11.471	107	179938	45.942	ug/l	99
64) Tetrachloroethene	11.106	164	180084	48.146	ug/l	97
65) Chlorobenzene	11.894	112	481213	44.108	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.965	131	186809	46.229	ug/l	100
67) Ethyl Benzene	11.965	91	865400	46.999	ug/l	100
68) m/p-Xylenes	12.071	106	688771	94.509	ug/l	99
69) o-Xylene	12.400	106	338574	46.945	ug/l	100
70) Styrene	12.412	104	558924	48.741	ug/l	99
71) Bromoform	12.582	173	142704	48.532	ug/l #	99
73) Isopropylbenzene	12.700	105	886254	47.187	ug/l	100
74) N-amyl acetate	12.494	43	269994	49.001	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.941	83	246842	42.514	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	219415m	41.881	ug/l	
77) Bromobenzene	12.982	156	211758	44.551	ug/l	99
78) n-propylbenzene	13.035	91	999457	48.248	ug/l	99
79) 2-Chlorotoluene	13.129	91	594476	45.973	ug/l	100
80) 1,3,5-Trimethylbenzene	13.176	105	748541	47.619	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	70072	48.143	ug/l	97
82) 4-Chlorotoluene	13.129	91	594476	45.973	ug/l	100
83) tert-Butylbenzene	13.441	119	696661	50.266	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	744208	47.723	ug/l	99
85) sec-Butylbenzene	13.618	105	949866	49.655	ug/l	99
86) p-Isopropyltoluene	13.729	119	798885	50.190	ug/l	100
87) 1,3-Dichlorobenzene	13.735	146	395107	45.000	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	391507	43.649	ug/l	100
89) n-Butylbenzene	14.059	91	629523	48.819	ug/l	100
90) Hexachloroethane	14.335	117	134261	47.154	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	395099	45.685	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.723	75	44797	43.482	ug/l	94
93) 1,2,4-Trichlorobenzene	15.394	180	199117	46.505	ug/l	99
94) Hexachlorobutadiene	15.506	225	105970	45.364	ug/l	99
95) Naphthalene	15.641	128	573904	40.984	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	189351	47.316	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
Data File : VN076258.D
Acq On : 13 Jan 2023 19:03
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023

Quant Time: Jan 16 01:15:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
Data File : VN076258.D
Acq On : 13 Jan 2023 19:03
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023

Quant Time: Jan 16 01:15:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1050
Delta R.T. -0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC






Tgt Ion:168 Resp: 377960
Ion Ratio Lower Upper
168 100
99 53.6 44.2 66.4

Manual Integrations

APPROVED




Reviewed By :John Carlone 01/16/2023

Supervised By :Mahesh Dadoda 01/16/2023



#2
Dichlorodifluoromethane
Concen: 56.443 ug/l
RT: 2.136 min Scan# 14
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Tgt Ion: 85 Resp: 151472
Ion Ratio Lower Upper
85 100
87 32.2 16.3 48.9









#3
Chloromethane
Concen: 43.017 ug/l
RT: 2.377 min Scan# 50
Delta R.T. 0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 50 Resp: 143549
Ion Ratio Lower Upper
50 100
52 33.5 27.7 41.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#4
Vinyl Chloride
Concen: 46.422 ug/l
RT: 2.524 min Scan# 80
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Tgt Ion: 62 Resp: 184788
Ion Ratio Lower Upper
62 100
64 33.4 24.9 37.3









#5
Bromomethane
Concen: 44.439 ug/l
RT: 2.948 min Scan# 151
Delta R.T. 0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 94 Resp: 150330
Ion Ratio Lower Upper
94 100
96 91.4 73.8 110.6




Manual Integrations
APPROVED

Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#6
Chloroethane
Concen: 47.304 ug/l
RT: 3.118 min Scan# 181
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03


Tgt Ion: 64 Resp: 139442
Ion Ratio Lower Upper
64 100
66 32.7 24.6 36.8





#7
 Trichlorofluoromethane
 Concen: 44.219 ug/l
 RT: 3.506 min Scan# 246
 Delta R.T. 0.006 min
 Lab File: VN076258.D
 Acq: 13 Jan 2023 19:03




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC



Tgt Ion:101 Resp: 26709
 Ion Ratio Lower Upper
 101 100
 103 66.8 53.4 80.2




Manual Integrations
 APPROVED


Reviewed By :John Carlone 01/16/2023
 Supervised By :Mahesh Dadoda 01/16/2023



#8
 Diethyl Ether
 Concen: 44.168 ug/l
 RT: 3.971 min Scan# 326
 Delta R.T. 0.000 min
 Lab File: VN076258.D
 Acq: 13 Jan 2023 19:03

Tgt Ion: 74 Resp: 97441
 Ion Ratio Lower Upper
 74 100
 45 75.0 37.8 113.4









#9
1,1,2-Trichlorotrifluoroethane
Concen: 44.613 ug/l
RT: 4.383 min Scan# 396
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 153874
Ion Ratio Lower Upper
101 100
85 45.2 35.3 52.9
151 81.0 65.6 98.4




Manual Integrations
APPROVED

Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#10
Methyl Iodide
Concen: 42.793 ug/l
RT: 4.600 min Scan# 433
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03


Tgt Ion:142 Resp: 239531
Ion Ratio Lower Upper
142 100
127 44.6 34.7 52.1
141 14.4 11.4 17.2





#11
Tert butyl alcohol
Concen: 212.654 ug/l
RT: 5.536 min Scan# 591
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 59 Resp: 12372
Ion Ratio Lower Upper
59 100
57 9.7 8.2 12.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#12
1,1-Dichloroethene
Concen: 42.757 ug/l
RT: 4.353 min Scan# 391
Delta R.T. 0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Tgt Ion: 96 Resp: 143228
Ion Ratio Lower Upper
96 100
61 143.5 113.3 169.9
98 61.7 50.9 76.3





#13
Acrolein
Concen: 225.151 ug/l
RT: 4.195 min Scan# 30
Delta R.T. 0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 56 Resp: 160599
Ion Ratio Lower Upper
56 100
55 68.8 54.8 82.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#14
Allyl chloride
Concen: 43.081 ug/l
RT: 5.036 min Scan# 507
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Tgt Ion: 41 Resp: 168700
Ion Ratio Lower Upper
41 100
39 84.0 58.3 87.5
76 49.6 35.8 53.6





#15
Acrylonitrile
Concen: 220.804 ug/l
RT: 5.730 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 53 Resp: 350584
Ion Ratio Lower Upper
53 100
52 81.8 65.4 98.0
51 35.5 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#16
Acetone
Concen: 194.436 ug/l
RT: 4.442 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03


Tgt Ion: 43 Resp: 275601
Ion Ratio Lower Upper
43 100
58 34.2 28.6 43.0





#17
Carbon Disulfide
Concen: 37.862 ug/l
RT: 4.724 min Scan# 454
Delta R.T. 0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 76 Resp: 307080
Ion Ratio Lower Upper
76 100
78 8.5 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#18
Methyl Acetate
Concen: 47.403 ug/l
RT: 5.042 min Scan# 508
Delta R.T. 0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Tgt Ion: 43 Resp: 198609
Ion Ratio Lower Upper
43 100
74 29.9 23.6 35.4








#19
Methyl tert-butyl Ether
Concen: 48.140 ug/l
RT: 5.806 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 73 Resp: 53104
Ion Ratio Lower Upper
73 100
57 21.0 17.1 25.7




Manual Integrations
APPROVED

Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#20
Methylene Chloride
Concen: 43.876 ug/l
RT: 5.289 min Scan# 550
Delta R.T. -0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03


Tgt Ion: 84 Resp: 166666
Ion Ratio Lower Upper
84 100
49 108.3 81.8 122.6
51 34.7 26.7 40.1
86 62.6 51.5 77.3





#21
trans-1,2-Dichloroethene
Concen: 41.550 ug/l
RT: 5.795 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 96 Resp: 152430
Ion Ratio Lower Upper
96 100
61 122.9 98.1 147.1
98 66.8 51.4 77.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#22
Diisopropyl ether
Concen: 45.845 ug/l
RT: 6.683 min Scan# 787
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03



Tgt Ion: 45 Resp: 436112
Ion Ratio Lower Upper
45 100
43 56.4 44.2 66.4
87 36.9 28.3 42.5
59 13.5 11.1 16.7








#23
 Vinyl Acetate
 Concen: 221.873 ug/l
 RT: 6.612 min Scan# 775
 Delta R.T. 0.000 min
 Lab File: VN076258.D
 Acq: 13 Jan 2023 19:03

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion: 43 Resp: 1420609
 Ion Ratio Lower Upper
 43 100
 86 14.0 11.4 17.2




Manual Integrations
 APPROVED


Reviewed By :John Carlone 01/16/2023
 Supervised By :Mahesh Dadoda 01/16/2023



#24
 1,1-Dichloroethane
 Concen: 44.096 ug/l
 RT: 6.577 min Scan# 769
 Delta R.T. 0.000 min
 Lab File: VN076258.D
 Acq: 13 Jan 2023 19:03


Tgt Ion: 63 Resp: 279754
 Ion Ratio Lower Upper
 63 100
 98 6.7 3.9 11.5
 100 4.7 2.5 7.6





#25
2-Butanone
Concen: 218.201 ug/l
RT: 7.489 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 43305
Ion Ratio Lower Upper
43 100
72 31.8 28.0 42.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#26
2,2-Dichloropropane
Concen: 43.288 ug/l
RT: 7.500 min Scan# 926
Delta R.T. 0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Tgt Ion: 77 Resp: 232571
Ion Ratio Lower Upper
77 100
97 23.8 12.7 38.0





#27
 cis-1,2-Dichloroethene
 Concen: 43.802 ug/l
 RT: 7.494 min Scan# 91
 Delta R.T. 0.000 min
 Lab File: VN076258.D
 Acq: 13 Jan 2023 19:03




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC



Tgt Ion: 96 Resp: 19218
 Ion Ratio Lower Upper
 96 100
 61 127.6 0.0 247.8
 98 63.8 0.0 128.4




Manual Integrations
 APPROVED

Reviewed By :John Carlone 01/16/2023
 Supervised By :Mahesh Dadoda 01/16/2023



#28
 Bromochloromethane
 Concen: 40.240 ug/l
 RT: 7.818 min Scan# 980
 Delta R.T. -0.006 min
 Lab File: VN076258.D
 Acq: 13 Jan 2023 19:03


Tgt Ion: 49 Resp: 111894
 Ion Ratio Lower Upper
 49 100
 129 1.6 0.0 5.0
 130 94.1 77.0 115.4





#29
Tetrahydrofuran
Concen: 232.094 ug/l
RT: 7.847 min Scan# 984
Delta R.T. 0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 42 Resp: 288664
Ion Ratio Lower Upper
42 100
72 53.3 44.5 66.7
71 50.3 41.0 61.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#30
Chloroform
Concen: 43.474 ug/l
RT: 7.971 min Scan# 1006
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Tgt Ion: 83 Resp: 310534
Ion Ratio Lower Upper
83 100
85 65.6 49.1 73.7





#31
Cyclohexane
Concen: 40.833 ug/l
RT: 8.259 min Scan# 1055
Delta R.T. -0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 56 Resp: 240462
Ion Ratio Lower Upper
56 100
69 34.2 24.7 37.1
84 98.7 76.5 114.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#32
1,1,1-Trichloroethane
Concen: 44.133 ug/l
RT: 8.171 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03



Tgt Ion: 97 Resp: 279479
Ion Ratio Lower Upper
97 100
99 63.4 50.2 75.2
61 45.2 34.5 51.7





#33

1,2-Dichloroethane-d4

Concen: 45.069 ug/l

RT: 8.583 min Scan# 1110

Delta R.T. 0.000 min

Lab File: VN076258.D

Acq: 13 Jan 2023 19:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 185310

Ion Ratio Lower Upper

65 100





67 52.1 0.0 103.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/16/2023

Supervised By :Mahesh Dadoda 01/16/2023



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1199

Delta R.T. 0.000 min

Lab File: VN076258.D

Acq: 13 Jan 2023 19:03




Tgt Ion:114 Resp: 596498

Ion Ratio Lower Upper

114 100

63 18.2 0.0 34.4

88 14.3 0.0 28.8





#35

Dibromofluoromethane

Concen: 46.815 ug/l

RT: 8.171 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VN076258.D

Acq: 13 Jan 2023 19:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion:113 Resp: 172934





Ion	Ratio	Lower	Upper
113	100		
111	101.8	83.4	125.2
192	20.0	16.1	24.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/16/2023

Supervised By :Mahesh Dadoda 01/16/2023



#36

1,1-Dichloropropene

Concen: 44.311 ug/l

RT: 8.377 min Scan# 1075




Delta R.T. 0.000 min


Lab File: VN076258.D

Acq: 13 Jan 2023 19:03

Tgt Ion: 75 Resp: 217439

Ion	Ratio	Lower	Upper
75	100		
110	37.7	18.8	56.4
77	30.5	24.1	36.1









#37
Ethyl Acetate
Concen: 46.540 ug/l
RT: 7.565 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 18344
Ion Ratio Lower Upper
43 100
61 14.8 13.0 19.4
70 11.8 9.4 14.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#38
Carbon Tetrachloride
Concen: 45.298 ug/l
RT: 8.371 min Scan# 1074
Delta R.T. 0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Tgt Ion:117 Resp: 248787
Ion Ratio Lower Upper
117 100
119 94.1 76.1 114.1
121 30.1 24.5 36.7





#39

Methylcyclohexane

Concen: 45.720 ug/l

RT: 9.606 min Scan# 1115

Delta R.T. 0.000 min

Lab File: VN076258.D

Acq: 13 Jan 2023 19:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 27066

Ion Ratio Lower Upper

83 100

55 64.7 51.4 77.0





98 49.5 39.8 59.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/16/2023

Supervised By :Mahesh Dadoda 01/16/2023



#40

Benzene

Concen: 44.146 ug/l

RT: 8.612 min Scan# 1115

Delta R.T. 0.000 min

Lab File: VN076258.D




Acq: 13 Jan 2023 19:03

Tgt Ion: 78 Resp: 668387

Ion Ratio Lower Upper

78 100


77 22.7 18.5 27.7





#41
Methacrylonitrile
Concen: 48.188 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 41 Resp: 98729
Ion Ratio Lower Upper
41 100
39 56.3 43.1 64.7
67 87.7 70.7 106.1
52 36.0 26.7 40.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#42
1,2-Dichloroethane
Concen: 45.228 ug/l
RT: 8.677 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Tgt Ion: 62 Resp: 227652
Ion Ratio Lower Upper
62 100
98 10.2 0.0 20.8





#43
Isopropyl Acetate
Concen: 49.931 ug/l
RT: 8.694 min Scan# 1129
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 324658
Ion Ratio Lower Upper
43 100
61 29.1 24.2 36.4
87 15.9 13.3 19.9

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#44
Trichloroethene
Concen: 42.972 ug/l
RT: 9.359 min Scan# 1242
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03



Tgt Ion:130 Resp: 183287
Ion Ratio Lower Upper
130 100
95 97.3 0.0 195.4





#45

1,2-Dichloropropane

Concen: 44.636 ug/l

RT: 9.624 min Scan# 1287

Delta R.T. 0.000 min

Lab File: VN076258.D

Acq: 13 Jan 2023 19:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 16445

Ion Ratio Lower Upper

63 100





65 32.1 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/16/2023

Supervised By :Mahesh Dadoda 01/16/2023



#46

Dibromomethane

Concen: 43.894 ug/l

RT: 9.712 min Scan# 1302

Delta R.T. 0.000 min

Lab File: VN076258.D

Acq: 13 Jan 2023 19:03




Tgt Ion: 93 Resp: 119948


Ion Ratio Lower Upper

93 100

95 83.5 65.2 97.8

174 103.5 82.4 123.6





#47

Bromodichloromethane

Concen: 45.321 ug/l

RT: 9.894 min Scan# 1333

Delta R.T. 0.006 min

Lab File: VN076258.D

Acq: 13 Jan 2023 19:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 24286





Ion	Ratio	Lower	Upper
83	100		
85	63.2	51.5	77.3
127	9.2	6.6	10.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/16/2023

Supervised By :Mahesh Dadoda 01/16/2023



#48

Methyl methacrylate

Concen: 50.589 ug/l

RT: 9.682 min Scan# 1297




Delta R.T. 0.000 min


Lab File: VN076258.D

Acq: 13 Jan 2023 19:03

Tgt Ion: 41 Resp: 148148

Ion	Ratio	Lower	Upper
41	100		
69	100.0	81.9	122.9
39	67.1	51.6	77.4





#49

1,4-Dioxane

Concen: 863.138 ug/l

RT: 9.694 min Scan# 1299

Delta R.T. -0.006 min

Lab File: VN076258.D

Acq: 13 Jan 2023 19:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 88 Resp: 6214

Ion Ratio Lower Upper

88 100

43 26.1 17.1 25.7





58 61.5 47.8 71.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/16/2023

Supervised By :Mahesh Dadoda 01/16/2023



#50

Toluene-d8

Concen: 44.391 ug/l

RT: 10.571 min Scan# 1448

Delta R.T. 0.000 min

Lab File: VN076258.D




Acq: 13 Jan 2023 19:03


Tgt Ion: 98 Resp: 613105

Ion Ratio Lower Upper

98 100

100 64.5 52.1 78.1









#51
4-Methyl-2-Pentanone
Concen: 237.502 ug/l
RT: 10.447 min Scan# 1427
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 933892
Ion Ratio Lower Upper
43 100
58 43.8 35.5 53.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#52
Toluene
Concen: 45.545 ug/l
RT: 10.630 min Scan# 1458
Delta R.T. -0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Tgt Ion: 92 Resp: 449091
Ion Ratio Lower Upper
92 100
91 169.6 137.3 205.9









#53
t-1,3-Dichloropropene
Concen: 47.081 ug/l
RT: 10.841 min Scan# 1493
Delta R.T. 0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 75 Resp: 24697
Ion Ratio Lower Upper
75 100
77 30.6 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#54
cis-1,3-Dichloropropene
Concen: 47.149 ug/l
RT: 10.312 min Scan# 1404
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03


Tgt Ion: 75 Resp: 271557
Ion Ratio Lower Upper
75 100
77 29.7 25.8 38.8
39 42.7 33.4 50.2





#55
1,1,2-Trichloroethane
Concen: 45.991 ug/l
RT: 11.018 min Scan# 1524
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 97 Resp: 177990
Ion Ratio Lower Upper
97 100
83 86.2 65.0 97.4
85 53.1 43.0 64.6
99 63.6 50.2 75.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#56
Ethyl methacrylate
Concen: 49.237 ug/l
RT: 10.877 min Scan# 1500
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Tgt Ion: 69 Resp: 253114
Ion Ratio Lower Upper
69 100
41 55.6 44.2 66.4
39 36.5 26.4 39.6





#57
1,3-Dichloropropane
Concen: 45.683 ug/l
RT: 11.165 min Scan# 1549
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 76 Resp: 290018
Ion Ratio Lower Upper
76 100
78 32.0 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#58
2-Chloroethyl Vinyl ether
Concen: 250.203 ug/l
RT: 10.165 min Scan# 1379
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03


Tgt Ion: 63 Resp: 409515
Ion Ratio Lower Upper
63 100
106 33.0 26.4 39.6





#59
2-Hexanone
Concen: 235.763 ug/l
RT: 11.200 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 676074
Ion Ratio Lower Upper
43 100
58 61.0 30.6 92.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#60
Dibromochloromethane
Concen: 46.426 ug/l
RT: 11.359 min Scan# 1582
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03


Tgt Ion:129 Resp: 199825
Ion Ratio Lower Upper
129 100
127 77.3 38.3 114.9





#61
1,2-Dibromoethane
Concen: 45.942 ug/l
RT: 11.471 min Scan# 1601
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:107 Resp: 179938
Ion Ratio Lower Upper
107 100
109 94.1 74.8 112.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#62
4-Bromofluorobenzene
Concen: 46.935 ug/l
RT: 12.853 min Scan# 1836
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Tgt Ion: 95 Resp: 214705
Ion Ratio Lower Upper
95 100
174 86.8 0.0 171.6
176 84.1 0.0 169.2





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:117 Resp: 531544
Ion Ratio Lower Upper
117 100
82 53.1 41.3 61.9
119 30.8 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#64
Tetrachloroethene
Concen: 48.146 ug/l
RT: 11.106 min Scan# 1539
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03


Tgt Ion:164 Resp: 180084
Ion Ratio Lower Upper
164 100
166 130.6 101.9 152.9
129 91.4 74.2 111.2
131 86.9 72.4 108.6





#65
Chlorobenzene
Concen: 44.108 ug/l
RT: 11.894 min Scan# 1673
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:112 Resp: 48121
Ion Ratio Lower Upper
112 100
114 32.4 24.7 37.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#66
1,1,1,2-Tetrachloroethane
Concen: 46.229 ug/l
RT: 11.965 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Tgt Ion:131 Resp: 186809
Ion Ratio Lower Upper
131 100
133 94.7 47.5 142.5
119 64.0 31.9 95.7





#67
Ethyl Benzene
Concen: 46.999 ug/l
RT: 11.965 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 91 Resp: 865400
Ion Ratio Lower Upper
91 100
106 32.1 25.7 38.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#68
m/p-Xylenes
Concen: 94.509 ug/l
RT: 12.071 min Scan# 1703
Delta R.T. -0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03



Tgt Ion:106 Resp: 688771
Ion Ratio Lower Upper
106 100
91 194.3 156.6 235.0









#69
o-Xylene
Concen: 46.945 ug/l
RT: 12.400 min Scan# 1759
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 338574
Ion Ratio Lower Upper
106 100
91 208.4 103.8 311.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#70
Styrene
Concen: 48.741 ug/l
RT: 12.412 min Scan# 1761
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03


Tgt Ion:104 Resp: 558924
Ion Ratio Lower Upper
104 100
78 46.5 37.5 56.3
103 54.1 43.5 65.3





#71
Bromoform
Concen: 48.532 ug/l
RT: 12.582 min Scan# 1790
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:173 Resp: 142704
Ion Ratio Lower Upper
173 100
175 47.8 24.3 72.8
254 0.1 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 1996
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Tgt Ion:152 Resp: 259903
Ion Ratio Lower Upper
152 100
115 57.6 29.1 87.4
150 171.4 0.0 347.0





#73
Isopropylbenzene
Concen: 47.187 ug/l
RT: 12.700 min Scan# 1810
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:105 Resp: 886254
Ion Ratio Lower Upper
105 100
120 26.8 13.4 40.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#74
N-ethyl acetate
Concen: 49.001 ug/l
RT: 12.494 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Tgt Ion: 43 Resp: 269994
Ion Ratio Lower Upper
43 100
70 50.4 41.4 62.0
55 26.0 21.3 31.9
61 26.9 21.2 31.8





#75
1,1,2,2-Tetrachloroethane
Concen: 42.514 ug/l
RT: 12.941 min Scan# 1851
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 24684
Ion Ratio Lower Upper
83 100
131 11.0 5.6 16.8
85 65.2 32.6 98.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#76
1,2,3-Trichloropropane
Concen: 41.881 ug/l m
RT: 12.994 min Scan# 1860
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03


Tgt Ion: 75 Resp: 219415
Ion Ratio Lower Upper
75 100
77 178.1 82.6 247.8





#77
Bromobenzene
Concen: 44.551 ug/l
RT: 12.982 min Scan# 1858
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:156 Resp: 211758
Ion Ratio Lower Upper
156 100
77 184.5 93.8 281.3
158 97.8 48.9 146.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#78
n-propylbenzene
Concen: 48.248 ug/l
RT: 13.035 min Scan# 1867
Delta R.T. -0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Tgt Ion: 91 Resp: 999457
Ion Ratio Lower Upper
91 100
120 23.7 12.0 36.1









#79
2-Chlorotoluene
Concen: 45.973 ug/l
RT: 13.129 min Scan# 1883
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 594470
Ion Ratio Lower Upper
91 100
126 34.8 17.5 52.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#80
1,3,5-Trimethylbenzene
Concen: 47.619 ug/l
RT: 13.176 min Scan# 1891
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Tgt Ion:105 Resp: 748541
Ion Ratio Lower Upper
105 100
120 50.0 24.7 74.1









#81
trans-1,4-Dichloro-2-butene
Concen: 48.143 ug/l
RT: 12.735 min Scan# 1816
Delta R.T. -0.006 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 75 Resp: 70072
Ion Ratio Lower Upper
75 100
53 95.4 73.8 110.6
89 48.5 39.5 59.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#82
4-Chlorotoluene
Concen: 45.973 ug/l
RT: 13.129 min Scan# 1883
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Tgt Ion: 91 Resp: 594476
Ion Ratio Lower Upper
91 100
126 34.8 17.5 52.5









#83
tert-Butylbenzene
Concen: 50.266 ug/l
RT: 13.441 min Scan# 1936
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:119 Resp: 69666
Ion Ratio Lower Upper
119 100
91 62.1 31.9 95.5
134 26.2 13.8 41.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#84
1,2,4-Trimethylbenzene
Concen: 47.723 ug/l
RT: 13.482 min Scan# 1943
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03


Tgt Ion:105 Resp: 744208
Ion Ratio Lower Upper
105 100
120 46.1 23.3 69.8





#85
 sec-Butylbenzene
 Concen: 49.655 ug/l
 RT: 13.618 min Scan# 1966
 Delta R.T. 0.000 min
 Lab File: VN076258.D
 Acq: 13 Jan 2023 19:03



Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion:105 Resp: 949860
 Ion Ratio Lower Upper
 105 100
 134 20.5 10.1 30.1

Manual Integrations
 APPROVED



Reviewed By :John Carlone 01/16/2023
 Supervised By :Mahesh Dadoda 01/16/2023




#86
 p-Isopropyltoluene
 Concen: 50.190 ug/l
 RT: 13.729 min Scan# 1985
 Delta R.T. 0.000 min
 Lab File: VN076258.D
 Acq: 13 Jan 2023 19:03




Tgt Ion:119 Resp: 798885
 Ion Ratio Lower Upper
 119 100
 134 26.1 13.0 39.0
 91 22.7 11.2 33.6





#87
1,3-Dichlorobenzene
Concen: 45.000 ug/l
RT: 13.735 min Scan# 1986
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 395107
Ion Ratio Lower Upper
146 100
111 39.6 19.7 59.0
148 63.4 31.4 94.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#88
1,4-Dichlorobenzene
Concen: 43.649 ug/l
RT: 13.812 min Scan# 1999
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Tgt Ion:146 Resp: 391507
Ion Ratio Lower Upper
146 100
111 38.6 19.1 57.1
148 64.3 32.0 96.2





#89
n-Butylbenzene
Concen: 48.819 ug/l
RT: 14.059 min Scan# 2041
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 62952
Ion Ratio Lower Upper
91 100
92 52.9 26.4 79.2
134 27.1 13.4 40.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#90
Hexachloroethane
Concen: 47.154 ug/l
RT: 14.335 min Scan# 2088
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03


Tgt Ion:117 Resp: 134261
Ion Ratio Lower Upper
117 100
201 76.7 38.0 114.1





#91
1,2-Dichlorobenzene
Concen: 45.685 ug/l
RT: 14.106 min Scan# 2049
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 395099
Ion Ratio Lower Upper
146 100
111 40.4 20.3 61.0
148 63.3 31.7 95.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#92
1,2-Dibromo-3-Chloropropane
Concen: 43.482 ug/l
RT: 14.723 min Scan# 2154
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Tgt Ion: 75 Resp: 44797
Ion Ratio Lower Upper
75 100
155 93.5 49.3 147.8
157 118.9 63.7 191.3





#93
1,2,4-Trichlorobenzene
Concen: 46.505 ug/l
RT: 15.394 min Scan# 2268
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:180 Resp: 19911
Ion Ratio Lower Upper
180 100
182 93.7 47.5 142.5
145 30.8 15.3 45.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023




#94
Hexachlorobutadiene
Concen: 45.364 ug/l
RT: 15.506 min Scan# 2287
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Tgt Ion:225 Resp: 105970
Ion Ratio Lower Upper
225 100
223 62.0 31.5 94.5
227 62.9 31.6 95.0





#95
Naphthalene
Concen: 40.984 ug/l
RT: 15.641 min Scan# 2310
Delta R.T. -0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:128 Resp: 573904
Ion Ratio Lower Upper
128 100
127 12.8 10.4 15.6
129 11.0 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/16/2023
Supervised By :Mahesh Dadoda 01/16/2023



#96
1,2,3-Trichlorobenzene
Concen: 47.316 ug/l
RT: 15.841 min Scan# 2344
Delta R.T. 0.000 min
Lab File: VN076258.D
Acq: 13 Jan 2023 19:03



Tgt Ion:180 Resp: 189351
Ion Ratio Lower Upper
180 100
182 95.5 47.0 141.0
145 32.4 16.0 47.9

