

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011325\
 Data File : VN085431.D
 Acq On : 13 Jan 2025 17:06
 Operator : JC\MD
 Sample : Q1062-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VPB190A-HYD-20250107

Quant Time: Jan 13 17:22:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 28 01:36:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	171762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	321172	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	285304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	114295	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	144844	59.124	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	118.240%	
35) Dibromofluoromethane	8.165	113	109949	54.874	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	109.740%	
50) Toluene-d8	10.565	98	353662	48.101	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	96.200%	
62) 4-Bromofluorobenzene	12.847	95	117799	48.242	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	96.480%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011325\
Data File : VN085431.D
Acq On : 13 Jan 2025 17:06
Operator : JC\MD
Sample : Q1062-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VPB190A-HYD-20250107

Quant Time: Jan 13 17:22:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
Quant Title : SW846 8260
QLast Update : Sat Dec 28 01:36:16 2024
Response via : Initial Calibration

