

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011325\
 Data File : VN085433.D
 Acq On : 13 Jan 2025 17:54
 Operator : JC\MD
 Sample : Q1062-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-190A-EB-20250107

Quant Time: Jan 13 18:11:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 28 01:36:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	177659	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	334626	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	294528	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	115338	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	146550	57.835	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 115.680%			
35) Dibromofluoromethane	8.165	113	108928	52.179	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.360%			
50) Toluene-d8	10.565	98	372474	48.623	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.240%			
62) 4-Bromofluorobenzene	12.847	95	121568	47.784	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 95.560%			
Target Compounds						
29) Tetrahydrofuran	7.841	42	5374	6.454	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011325\
Data File : VN085433.D
Acq On : 13 Jan 2025 17:54
Operator : JC\MD
Sample : Q1062-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-190A-EB-20250107

Quant Time: Jan 13 18:11:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
Quant Title : SW846 8260
QLast Update : Sat Dec 28 01:36:16 2024
Response via : Initial Calibration

