

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011419\
 Data File : VN053380.D
 Acq On : 14 Jan 2019 10:11
 Operator : JC/SP
 Sample : VN0114WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0114WBL01

Quant Time: Jan 14 23:12:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 07 09:34:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	876044	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1351348	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1267354	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	542526	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	433845	47.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.20%	
35) Dibromofluoromethane	7.59	113	399062	51.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.74%	
50) Toluene-d8	10.09	98	1677350	50.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.08%	
62) 4-Bromofluorobenzene	12.40	95	580948	50.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.00%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011419\
Data File : VN053380.D
Acq On : 14 Jan 2019 10:11
Operator : JC/SP
Sample : VN0114WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN0114WBL01

Quant Time: Jan 14 23:12:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 07 09:34:27 2019
Response via : Initial Calibration

