

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN011420\
 Data File : VN059700.D
 Acq On : 14 Jan 2020 12:03
 Operator : JC/MD
 Sample : PB126072TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB126072TB

Quant Time: Jan 14 22:16:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 14 08:58:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	187243	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	278367	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	237243	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	103578	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	96963	46.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.26%	
35) Dibromofluoromethane	7.57	113	85450	51.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.96%	
50) Toluene-d8	10.08	98	312647	48.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.22%	
62) 4-Bromofluorobenzene	12.40	95	109764	47.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.68%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.79	43	4491	Below Cal		97
17) Carbon Disulfide	4.03	76	1270	0.260	ug/l #	75
18) Methyl Acetate	4.30	43	1497	0.616	ug/l #	58
20) Methylene Chloride	4.52	84	3115	0.371	ug/l	94
25) 2-Butanone	6.81	43	1451	1.232	ug/l #	50
31) Cyclohexane	7.65	56	3751	1.219	ug/l #	12
36) 1,1-Dichloropropene	7.64	75	12673	5.228	ug/l #	50
38) Carbon Tetrachloride	7.64	117	17045	6.464	ug/l #	15
43) Isopropyl Acetate	8.14	43	43940	11.236	ug/l #	91
45) 1,2-Dichloropropane	9.25	63	957	0.535	ug/l #	44
48) Methyl methacrylate	9.16	41	2640	1.509	ug/l #	13
51) 4-Methyl-2-Pentanone	9.97	43	3382	1.496	ug/l	97
52) Toluene	10.14	92	1746	0.382	ug/l	90
54) cis-1,3-Dichloropropene	9.89	75	21527	7.205	ug/l #	57
57) 1,3-Dichloropropane	10.76	76	3488	1.215	ug/l #	42
59) 2-Hexanone	10.75	43	47001	28.221	ug/l #	55
71) Bromoform	12.40	173	324	0.224	ug/l #	1
74) N-amyl acetate	12.03	43	3293	1.117	ug/l #	45
76) 1,2,3-Trichloropropane	12.40	75	55632	31.915	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.73	105	4135	0.647	ug/l	97
81) trans-1,4-Dichloro-2-buten	12.40	75	55632	68.538	ug/l #	11
88) 1,4-Dichlorobenzene	13.36	146	670	0.201	ug/l #	24
92) 1,2-Dibromo-3-Chloropropan	14.27	75	186	Below Cal	#	6
93) 1,2,4-Trichlorobenzene	14.91	180	1094	0.551	ug/l	87
94) Hexachlorobutadiene	15.01	225	306	0.280	ug/l	76
95) Naphthalene	15.13	128	4157	0.736	ug/l	99
96) 1,2,3-Trichlorobenzene	15.31	180	1359	0.703	ug/l	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN011420\
Data File : VN059700.D
Acq On : 14 Jan 2020 12:03
Operator : JC/MD
Sample : PB126072TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB126072TB

Quant Time: Jan 14 22:16:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 14 08:58:21 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1826

Delta R.T. 0.00 min

Lab File: VN059700.D

Acq: 14 Jan 2020 12:03

Instrument :

MSVOA_N

ClientSampleId :


PB126072TB

Tgt Ion:168 Resp: 187243

Ion Ratio Lower Upper




168 100

99 51.7 48.2 72.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#16

Acetone

Concen: Below Cal

RT: 3.79 min Scan# 627

Delta R.T. 0.01 min

Lab File: VN059700.D


Acq: 14 Jan 2020 12:03

Tgt Ion: 43 Resp: 4491

Ion Ratio Lower Upper



43 100


58 28.0 23.6 35.4



Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 58.00 (57.70 to 58.70): VNO









#17
Carbon Disulfide
Concen: 0.260 ug/l
RT: 4.03 min Scan# 703
Delta R.T. 0.00 min
Lab File: VN059700.D
Acq: 14 Jan 2020 12:03




Instrument :
MSVOA_N
ClientSampled :
PB126072TB


Tgt Ion: 76 Resp: 1270
Ion Ratio Lower Upper
76 100
78 0.0 7.4 11.0#



#18
Methyl Acetate
Concen: 0.616 ug/l
RT: 4.30 min Scan# 786
Delta R.T. 0.01 min
Lab File: VN059700.D
Acq: 14 Jan 2020 12:03

Tgt Ion: 43 Resp: 1497
Ion Ratio Lower Upper
43 100
74 1.9 17.5 26.3#









#20
Methylene Chloride
Concen: 0.371 ug/l
RT: 4.52 min Scan# 855
Delta R.T. 0.00 min
Lab File: VN059700.D
Acq: 14 Jan 2020 12:03




Instrument :
MSVOA_N
ClientSampled :
PB126072TB


Tgt Ion:	84	Resp:	3115
Ion	Ratio	Lower	Upper
84	100		
49	123.4	106.6	160.0
51	36.4	31.9	47.9
86	64.1	50.8	76.2



#25
2-Butanone
Concen: 1.232 ug/l
RT: 6.81 min Scan# 1568
Delta R.T. 0.02 min
Lab File: VN059700.D
Acq: 14 Jan 2020 12:03

Tgt Ion:	43	Resp:	1451
Ion	Ratio	Lower	Upper
43	100		
72	0.0	19.9	29.9#









#31
Cyclohexane
Concen: 1.219 ug/l
RT: 7.65 min Scan# 1827
Delta R.T. 0.02 min
Lab File: VN059700.D
Acq: 14 Jan 2020 12:03

Instrument :
MSVOA_N
ClientSampled :
PB126072TB

Tgt Ion: 56 Resp: 3751
Ion Ratio Lower Upper
56 100
69 145.1 26.5 39.7#
84 120.4 65.3 97.9#




Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0






#33
1,2-Dichloroethane-d4
Concen: 46.631 ug/l
RT: 8.00 min Scan# 1938
Delta R.T. 0.00 min
Lab File: VN059700.D
Acq: 14 Jan 2020 12:03

Tgt Ion: 65 Resp: 96963
Ion Ratio Lower Upper
65 100
67 51.4 0.0 103.0



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.57 min Scan# 2113

Delta R.T. 0.00 min

Lab File: VN059700.D

Acq: 14 Jan 2020 12:03

Instrument :

MSVOA_N

ClientSampleId :

PB126072TB


Tgt Ion:114 Resp: 278367

Ion Ratio Lower Upper

114 100

63 18.8 0.0 43.6




88 15.0 0.0 31.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 87.90 (87.60 to 88.60): VN0



#35

Dibromofluoromethane

Concen: 51.475 ug/l

RT: 7.57 min Scan# 1802

Delta R.T. 0.00 min

Lab File: VN059700.D

Acq: 14 Jan 2020 12:03


Tgt Ion:113 Resp: 85450

Ion Ratio Lower Upper

113 100

111 102.5 83.5 125.3



192 21.3 15.4 23.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 5.228 ug/l
 RT: 7.64 min Scan# 1826
 Delta R.T. -0.13 min
 Lab File: VN059700.D
 Acq: 14 Jan 2020 12:03




Instrument :
 MSVOA_N
 ClientSampleId :
 PB126072TB


Tgt Ion: 75 Resp: 12673
 Ion Ratio Lower Upper
 75 100
 110 8.0 16.6 49.7#
 77 0.0 25.1 37.7#



#38
 Carbon Tetrachloride
 Concen: 6.464 ug/l
 RT: 7.64 min Scan# 1826
 Delta R.T. -0.11 min
 Lab File: VN059700.D
 Acq: 14 Jan 2020 12:03

Tgt Ion: 117 Resp: 17045
 Ion Ratio Lower Upper
 117 100
 119 4.3 76.5 114.7#
 121 0.0 25.1 37.7#





#43

Isopropyl Acetate

Concen: 11.236 ug/l

RT: 8.14 min Scan# 1980

Delta R.T. 0.00 min

Lab File: VN059700.D

Acq: 14 Jan 2020 12:03

Instrument :

MSVOA_N

ClientSampled :

PB126072TB


Tgt Ion: 43 Resp: 43940

Ion Ratio Lower Upper

43 100

61 20.6 20.2 30.2




87 14.7 9.5 14.3#



Abundance Ion 43.00 (42.70 to 43.70): VN059700.D (-1887) (-)

Ion 61.00 (60.70 to 61.70): VN059700.D (-1887) (-)

Ion 87.00 (86.70 to 87.70): VN059700.D (-1887) (-)



#45

1,2-Dichloropropane

Concen: 0.535 ug/l

RT: 9.25 min Scan# 2326

Delta R.T. 0.15 min

Lab File: VN059700.D


Acq: 14 Jan 2020 12:03

Tgt Ion: 63 Resp: 957

Ion Ratio Lower Upper



63 100


65 0.0 24.9 37.3#



Abundance Ion 62.90 (62.60 to 63.60): VN059700.D (-2186) (-)

Ion 64.90 (64.60 to 65.60): VN059700.D (-2186) (-)









#48
Methyl methacrylate
Concen: 1.509 ug/l
RT: 9.16 min Scan# 2298
Delta R.T. -0.02 min
Lab File: VN059700.D
Acq: 14 Jan 2020 12:03




Instrument :
MSVOA_N
ClientSampleId :
PB126072TB


Tgt Ion: 41 Resp: 2640
Ion Ratio Lower Upper
41 100
69 0.0 64.4 96.6#
39 0.0 51.3 76.9#



#50
Toluene-d8
Concen: 48.610 ug/l
RT: 10.08 min Scan# 2583
Delta R.T. 0.00 min
Lab File: VN059700.D
Acq: 14 Jan 2020 12:03

Tgt Ion: 98 Resp: 312647
Ion Ratio Lower Upper
98 100
100 64.5 51.2 76.8









#51
 4-Methyl-2-Pentanone
 Concen: 1.496 ug/l
 RT: 9.97 min Scan# 2550
 Delta R.T. 0.00 min
 Lab File: VN059700.D
 Acq: 14 Jan 2020 12:03




Instrument :
 MSVOA_N
 ClientSampled :
 PB126072TB


Tgt Ion: 43 Resp: 3382
 Ion Ratio Lower Upper
 43 100
 58 35.3 29.8 44.8



#52
 Toluene
 Concen: 0.382 ug/l
 RT: 10.14 min Scan# 2603
 Delta R.T. 0.00 min
 Lab File: VN059700.D
 Acq: 14 Jan 2020 12:03

Tgt Ion: 92 Resp: 1746
 Ion Ratio Lower Upper
 92 100
 91 189.1 140.4 210.6









#54
 cis-1,3-Dichloropropene
 Concen: 7.205 ug/l
 RT: 9.89 min Scan# 2524
 Delta R.T. 0.06 min
 Lab File: VN059700.D
 Acq: 14 Jan 2020 12:03




Instrument :
 MSVOA_N
 ClientSampleId :
 PB126072TB


Tgt Ion: 75 Resp: 21527
 Ion Ratio Lower Upper
 75 100
 77 0.0 25.1 37.7#
 39 31.0 46.1 69.1#



#57
 1,3-Dichloropropane
 Concen: 1.215 ug/l
 RT: 10.76 min Scan# 2794
 Delta R.T. 0.06 min
 Lab File: VN059700.D
 Acq: 14 Jan 2020 12:03

Tgt Ion: 76 Resp: 3488
 Ion Ratio Lower Upper
 76 100
 78 0.0 25.8 38.8#









#59
2-Hexanone
Concen: 28.221 ug/l
RT: 10.75 min Scan# 2793
Delta R.T. 0.01 min
Lab File: VN059700.D
Acq: 14 Jan 2020 12:03




Instrument :
MSVOA_N
ClientSampled :
PB126072TB


Tgt Ion: 43 Resp: 47001
Ion Ratio Lower Upper
43 100
58 20.2 25.8 77.3#



#62
4-Bromofluorobenzene
Concen: 47.340 ug/l
RT: 12.40 min Scan# 3304
Delta R.T. 0.00 min
Lab File: VN059700.D
Acq: 14 Jan 2020 12:03

Tgt Ion: 95 Resp: 109764
Ion Ratio Lower Upper
95 100
174 87.8 0.0 151.4
176 84.9 0.0 146.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.40 min Scan# 2994
Delta R.T. 0.00 min
Lab File: VN059700.D
Acq: 14 Jan 2020 12:03

Instrument :
MSVOA_N
ClientSampleId :
PB126072TB

Tgt Ion:117 Resp: 237243
Ion Ratio Lower Upper
117 100
82 52.7 45.9 68.9
119 32.4 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#71
Bromoform
Concen: 0.224 ug/l
RT: 12.40 min Scan# 3304
Delta R.T. 0.28 min
Lab File: VN059700.D
Acq: 14 Jan 2020 12:03

Tgt Ion:173 Resp: 324
Ion Ratio Lower Upper
173 100
175 1882.7 23.8 71.4#
254 0.0 0.0 0.0



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3597

Delta R.T. 0.00 min

Lab File: VN059700.D

Acq: 14 Jan 2020 12:03

Instrument :

MSVOA_N

ClientSampleId :

PB126072TB


Tgt Ion:152 Resp: 103578

Ion Ratio Lower Upper

152 100

115 57.8 30.3 90.9




150 157.5 0.0 345.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 1.117 ug/l

RT: 12.03 min Scan# 3189

Delta R.T. -0.04 min

Lab File: VN059700.D

Acq: 14 Jan 2020 12:03

Tgt Ion: 43 Resp: 3293


Ion Ratio Lower Upper

43 100

70 0.0 31.3 46.9#

55 0.0 19.8 29.6#

61 0.0 18.3 27.5#






Abundance Ion 43.00 (42.70 to 43.70): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 61.00 (60.70 to 61.70): V





#76

1,2,3-Trichloropropane

Concen: 31.915 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. -0.15 min

Lab File: VN059700.D

Acq: 14 Jan 2020 12:03

Instrument :

MSVOA_N

ClientSampleId :


PB126072TB

Tgt Ion: 75 Resp: 55632

Ion Ratio Lower Upper




75 100

77 1.2 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN059690.D (-3347) (-)

Ion 77.00 (76.70 to 77.70): VN059690.D (-3347) (-)



#80

1,3,5-Trimethylbenzene

Concen: 0.647 ug/l

RT: 12.73 min Scan# 3408

Delta R.T. 0.00 min

Lab File: VN059700.D


Acq: 14 Jan 2020 12:03

Tgt Ion: 105 Resp: 4135

Ion Ratio Lower Upper



105 100


120 50.6 24.1 72.4



Abundance Ion 105.00 (104.70 to 105.70): VN059690.D (-3390) (-)

Ion 120.00 (119.70 to 120.70): VN059690.D (-3390) (-)









#81
 trans-1,4-Dichloro-2-butene
 Concen: 68.538 ug/l
 RT: 12.40 min Scan# 3304
 Delta R.T. 0.10 min
 Lab File: VN059700.D
 Acq: 14 Jan 2020 12:03




Instrument :
 MSVOA_N
 ClientSampled :
 PB126072TB


Tgt Ion: 75 Resp: 55632
 Ion Ratio Lower Upper
 75 100
 53 0.0 78.1 117.1#
 89 0.0 35.4 53.2#



#88
 1,4-Dichlorobenzene
 Concen: 0.201 ug/l
 RT: 13.36 min Scan# 3603
 Delta R.T. 0.00 min
 Lab File: VN059700.D
 Acq: 14 Jan 2020 12:03

Tgt Ion: 146 Resp: 670
 Ion Ratio Lower Upper
 146 100
 111 0.0 20.8 62.2#
 148 127.6 31.8 95.4#









#92
 1,2-Dibromo-3-Chloropropane
 Concen: Below Cal
 RT: 14.27 min Scan# 3887
 Delta R.T. 0.01 min
 Lab File: VN059700.D
 Acq: 14 Jan 2020 12:03




Instrument :
 MSVOA_N
 ClientSampleId :
 PB126072TB


Tgt Ion: 75 Resp: 186
 Ion Ratio Lower Upper
 75 100
 155 0.0 37.0 110.9#
 157 0.0 48.4 145.2#



#93
 1,2,4-Trichlorobenzene
 Concen: 0.551 ug/l
 RT: 14.91 min Scan# 4086
 Delta R.T. 0.01 min
 Lab File: VN059700.D
 Acq: 14 Jan 2020 12:03

Tgt Ion: 180 Resp: 1094
 Ion Ratio Lower Upper
 180 100
 182 79.7 46.9 140.7
 145 35.5 15.0 45.1





#94

Hexachlorobutadiene

Concen: 0.280 ug/l

RT: 15.01 min Scan# 4117

Delta R.T. 0.00 min

Lab File: VN059700.D

Acq: 14 Jan 2020 12:03


Instrument :

MSVOA_N



ClientSampleId :

PB126072TB


Tgt Ion:	225	Resp:	306
Ion	Ratio	Lower	Upper
225	100		
223	80.7	31.8	95.3
227	44.1	32.1	96.5



Abundance Ion 224.70 (224.40 to 225.40): V
Ion 222.70 (222.40 to 223.40): V
Ion 226.70 (226.40 to 227.40): V



Time-->



#95

Naphthalene

Concen: 0.736 ug/l


RT: 15.13 min Scan# 4154

Delta R.T. 0.00 min



Lab File: VN059700.D

Acq: 14 Jan 2020 12:03


Tgt Ion:	128	Resp:	4157
Ion	Ratio	Lower	Upper
128	100		
127	13.5	10.4	15.6
129	11.4	8.6	12.8



Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V



Time-->



#96
 1,2,3-Trichlorobenzene
 Concen: 0.703 ug/l
 RT: 15.31 min Scan# 4211
 Delta R.T. 0.00 min
 Lab File: VN059700.D
 Acq: 14 Jan 2020 12:03

Instrument :
 MSVOA_N
 ClientSampleId :
 PB126072TB

Tgt Ion	Ratio	Lower	Upper
180	100		
182	97.7	47.4	142.3
145	47.5	16.3	48.9

