

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN011420\
 Data File : VN059707.D
 Acq On : 14 Jan 2020 16:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 14 23:09:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 14 08:58:21 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.534	0.595	-11.4	100	0.00
3 P	Chloromethane	0.572	0.561	1.9	98	0.00
4 C	Vinyl Chloride	0.655	0.680	-3.8#	97	0.00
5 T	Bromomethane	0.400	0.418	-4.5	103	0.00
6 T	Chloroethane	0.310	0.314	-1.3	101	0.00
7 T	Trichlorofluoromethane	0.787	0.808	-2.7	101	0.00
8 T	Diethyl Ether	0.276	0.279	-1.1	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.438	0.462	-5.5	103	0.00
10 T	Methyl Iodide	0.708	0.706	0.3	98	0.00
11 T	Tert butyl alcohol	0.093	0.080	14.0	84	0.00
12 CM	1,1-Dichloroethene	0.457	0.457	0.0	100	0.00
13 T	Acrolein	0.066	0.052	21.2	81	0.00
14 T	Allyl chloride	0.617	0.622	-0.8	99	0.00
15 T	Acrylonitrile	0.213	0.201	5.6	90	0.00
16 T	Acetone	0.224	0.174	22.3	88	0.00
17 T	Carbon Disulfide	1.303	1.283	1.5	98	0.00
18 T	Methyl Acetate	0.649	0.598	7.9	92	0.00
19 T	Methyl tert-butyl Ether	1.491	1.489	0.1	99	0.00
20 T	Methylene Chloride	0.557	0.497	10.8	100	0.00
21 T	trans-1,2-Dichloroethene	0.486	0.486	0.0	100	0.00
22 T	Diisopropyl ether	1.397	1.391	0.4	99	0.00
23 T	Vinyl Acetate	1.082	1.096	-1.3	106	0.00
24 P	1,1-Dichloroethane	0.854	0.853	0.1	100	0.00
25 T	2-Butanone	0.314	0.269	14.3	87	0.00
26 T	2,2-Dichloropropane	0.821	0.804	2.1	100	0.00
27 T	cis-1,2-Dichloroethene	0.559	0.564	-0.9	100	0.00
28 T	Bromochloromethane	0.341	0.325	4.7	96	0.00
29 T	Tetrahydrofuran	0.195	0.174	10.8	89	0.00
30 C	Chloroform	0.885	0.902	-1.9#	101	0.00
31 T	Cyclohexane	0.822	0.778	5.4	98	0.00
32 T	1,1,1-Trichloroethane	0.818	0.825	-0.9	99	0.00
33 S	1,2-Dichloroethane-d4	0.555	0.532	4.1	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.298	0.301	-1.0	98	0.00
36 T	1,1-Dichloropropene	0.435	0.441	-1.4	99	0.00
37 T	Ethyl Acetate	0.408	0.376	7.8	94	0.00
38 T	Carbon Tetrachloride	0.474	0.488	-3.0	103	0.00
39 T	Methylcyclohexane	0.537	0.548	-2.0	101	0.00
40 TM	Benzene	1.299	1.302	-0.2	101	0.00
41 T	Methacrylonitrile	0.201	0.178	11.4	89	0.00
42 TM	1,2-Dichloroethane	0.440	0.441	-0.2	99	0.00
43 T	Isopropyl Acetate	0.702	0.654	6.8	93	0.00
44 TM	Trichloroethene	0.354	0.373	-5.4	101	0.00
45 C	1,2-Dichloropropane	0.321	0.331	-3.1#	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN011420\
 Data File : VN059707.D
 Acq On : 14 Jan 2020 16:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 14 23:09:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 14 08:58:21 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.227	0.233	-2.6	101	0.00
47 T	Bromodichloromethane	0.457	0.474	-3.7	102	0.00
48 T	Methyl methacrylate	0.314	0.299	4.8	93	0.00
49 T	1,4-Dioxane	0.006	0.006	0.0	89	0.00
50 S	Toluene-d8	1.155	1.146	0.8	96	0.00
51 T	4-Methyl-2-Pentanone	0.406	0.369	9.1	90	0.00
52 CM	Toluene	0.820	0.842	-2.7#	101	0.00
53 T	t-1,3-Dichloropropene	0.510	0.530	-3.9	100	0.00
54 T	cis-1,3-Dichloropropene	0.537	0.569	-6.0	101	0.00
55 T	1,1,2-Trichloroethane	0.325	0.323	0.6	100	0.00
56 T	Ethyl methacrylate	0.496	0.497	-0.2	97	0.00
57 T	1,3-Dichloropropane	0.516	0.535	-3.7	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.150	0.169	-12.7	99	0.00
59 T	2-Hexanone	0.299	0.270	9.7	85	0.00
60 T	Dibromochloromethane	0.370	0.384	-3.8	101	0.00
61 T	1,2-Dibromoethane	0.340	0.350	-2.9	99	0.00
62 S	4-Bromofluorobenzene	0.416	0.416	0.0	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.408	0.357	12.5	81	0.00
65 PM	Chlorobenzene	0.983	1.012	-3.0	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.389	-2.4	103	0.00
67 C	Ethyl Benzene	1.784	1.814	-1.7#	102	0.00
68 T	m/p-Xylenes	0.674	0.696	-3.3	102	0.00
69 T	o-Xylene	0.660	0.670	-1.5	102	0.00
70 T	Styrene	1.104	1.140	-3.3	100	0.00
71 P	Bromoform	0.304	0.317	-4.3	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.667	3.657	0.3	102	0.00
74 T	N-amyl acetate	1.423	1.322	7.1	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.063	1.012	4.8	97	0.00
76 T	1,2,3-Trichloropropane	0.841	0.962	-14.4	113	0.00
77 T	Bromobenzene	0.929	0.928	0.1	102	0.00
78 T	n-propylbenzene	4.116	4.175	-1.4	103	0.00
79 T	2-Chlorotoluene	2.448	2.412	1.5	102	0.00
80 T	1,3,5-Trimethylbenzene	3.084	3.086	-0.1	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.392	0.371	5.4	94	0.00
82 T	4-Chlorotoluene	2.456	2.493	-1.5	103	0.00
83 T	tert-Butylbenzene	2.667	2.660	0.3	102	0.00
84 T	1,2,4-Trimethylbenzene	3.051	3.087	-1.2	103	0.00
85 T	sec-Butylbenzene	3.504	3.580	-2.2	103	0.00
86 T	p-Isopropyltoluene	3.243	3.286	-1.3	104	0.00
87 T	1,3-Dichlorobenzene	1.613	1.670	-3.5	105	0.00
88 T	1,4-Dichlorobenzene	1.609	1.642	-2.1	104	0.00
89 T	n-Butylbenzene	2.765	2.892	-4.6	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN011420\
Data File : VN059707.D
Acq On : 14 Jan 2020 16:15
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 14 23:09:11 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 14 08:58:21 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.607	0.611	-0.7	102	0.00
91 T	1,2-Dichlorobenzene	1.595	1.617	-1.4	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.211	17.9	93	0.00
93 T	1,2,4-Trichlorobenzene	0.958	1.025	-7.0	101	0.00
94 T	Hexachlorobutadiene	0.528	0.532	-0.8	102	0.00
95 T	Naphthalene	2.726	2.716	0.4	94	0.00
96 T	1,2,3-Trichlorobenzene	0.933	0.959	-2.8	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5