

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011421\
 Data File : VN065506.D
 Acq On : 14 Jan 2021 13:53
 Operator : JC/MD
 Sample : M1032-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 002-35TH-AVE(JAN)

Quant Time: Jan 14 14:29:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jan 14 04:12:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.158	128	40948	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.553	114	214291	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.380	117	189059	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.991	65	79319	31.01	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.37%
60) 4-Bromofluorobenzene	12.376	95	72187	25.21	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	84.03%
63) Toluene-d8	10.061	98	253879	30.68	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.27%
Target Compounds						
					Qvalue	
15) Acetone	3.791	58	174303	568.87	ug/l	92
16) Carbon Disulfide	4.007	76	3407	0.62	ug/l #	73
30) 2-Butanone	6.804	43	11154	9.08	ug/l #	92
58) 4-Methyl-2-Pentanone	9.955	43	16258	6.48	ug/l	95
62) Toluene	10.126	91	5493	0.56	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011421\
Data File : VN065506.D
Acq On : 14 Jan 2021 13:53
Operator : JC/MD
Sample : M1032-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(JAN)

Quant Time: Jan 14 14:29:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jan 14 04:12:08 2021
Response via : Initial Calibration

