

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011421\
 Data File : VN065509.D
 Acq On : 14 Jan 2021 15:07
 Operator : JC/MD
 Sample : M1102-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 55-ST-30

Quant Time: Jan 15 04:14:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jan 14 04:12:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.158	128	40583	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.553	114	210653	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.380	117	184572	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.994	65	78191	30.84	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.80%
60) 4-Bromofluorobenzene	12.376	95	68695	24.57	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	81.90%
63) Toluene-d8	10.061	98	246774	30.54	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.80%
Target Compounds						
15) Acetone	3.792	58	99564	327.87	ug/l	Qvalue 91
22) cis-1,2-Dichloroethene	6.788	96	17236	6.63	ug/l	88
30) 2-Butanone	6.804	43	282952	234.37	ug/l	98
35) Methacrylonitrile	7.180	41	192353	153.44	ug/l #	42

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011421\
Data File : VN065509.D
Acq On : 14 Jan 2021 15:07
Operator : JC/MD
Sample : M1102-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
55-ST-30

Quant Time: Jan 15 04:14:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jan 14 04:12:08 2021
Response via : Initial Calibration

