

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011421\
 Data File : VN065512.D
 Acq On : 14 Jan 2021 16:21
 Operator : JC/MD
 Sample : M1137-02 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DSN-002-SEWER-OUTLET-002

Quant Time: Jan 15 04:14:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jan 14 04:12:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.158	128	39770	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.553	114	213177	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.379	117	190450	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.994	65	78286	31.51	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	105.03%
60) 4-Bromofluorobenzene	12.376	95	80494	27.91	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.03%
63) Toluene-d8	10.061	98	247342	29.67	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.90%
Target Compounds						
15) Acetone	3.798	58	148927	500.45	ug/l	Qvalue 94
30) 2-Butanone	6.814	43	14027	11.48	ug/l	100
44) Isopropyl Acetate	8.135	43	20388	4.68	ug/l	99
82) p-Isopropyltoluene	13.264	119	18676	1.96	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011421\
Data File : VN065512.D
Acq On : 14 Jan 2021 16:21
Operator : JC/MD
Sample : M1137-02 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DSN-002-SEWER-OUTLET-002

Quant Time: Jan 15 04:14:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jan 14 04:12:08 2021
Response via : Initial Calibration

