

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011421\
 Data File : VN065501.D
 Acq On : 14 Jan 2021 11:10
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Jan 14 16:01:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jan 14 04:12:08 2021
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	101	0.00
2 M	Dichlorodifluoromethane	1.493	1.379	7.6	103	0.00
3 M	Chloromethane	1.848	1.662	10.1	99	0.00
4 M	Vinyl Chloride	2.036	1.829	10.2	102	0.00
5 M	Bromomethane	1.452	1.366	5.9	102	0.00
6 M	Chloroethane	1.324	1.191	10.0	101	0.00
7 M	Trichlorofluoromethane	2.712	2.468	9.0	100	0.00
8 T	Diethyl Ether	1.103	1.024	7.2	101	0.00
9	1,1,2-Trichlorotrifluoroeth	1.611	1.522	5.5	103	0.00
10 M	1,1-Dichloroethene	1.650	1.527	7.5	103	0.00
11	Methyl Iodide	2.604	2.315	11.1	99	0.00
12	Methyl Acetate	1.616	1.499	7.2	100	0.00
13 M	Acrolein	0.273	0.269	1.5	110	0.00
14 M	Acrylonitrile	0.732	0.678	7.4	100	0.00
15 M	Acetone	0.224	0.234	-4.5	111	0.00
16 M	Carbon Disulfide	4.011	3.581	10.7	102	0.00
17	Allyl chloride	2.339	2.093	10.5	102	0.00
18 M	Methylene Chloride	1.790	1.678	6.3	103	0.00
19 M	trans-1,2-Dichloroethene	1.660	1.524	8.2	102	0.00
20 T	Diisopropyl ether	4.899	4.519	7.8	102	0.00
21 M	1,1-Dichloroethane	2.927	2.693	8.0	101	0.00
22 M	cis-1,2-Dichloroethene	1.923	1.755	8.7	103	0.00
23 M	tert-Butyl Alcohol	0.226	0.214	5.3	101	0.00
24 M	Methyl tert-Butyl Ether	4.838	4.474	7.5	104	0.00
25 M	Chloroform	3.020	2.733	9.5	100	0.00
26	Cyclohexane	2.457	2.181	11.2	101	0.00
27 s	1,2-Dichloroethane-d4	1.874	1.862	0.6	98	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
29	1,1-Dichloropropene	0.426	0.390	8.5	102	0.00
30 M	2-Butanone	0.172	0.166	3.5	102	0.00
31	2,2-Dichloropropane	0.440	0.413	6.1	103	0.00
32 M	1,1,1-Trichloroethane	0.482	0.445	7.7	101	0.00
33 M	Carbon Tetrachloride	0.424	0.383	9.7	100	0.00
34 M	Benzene	1.312	1.211	7.7	101	0.00
35	Methacrylonitrile	0.179	0.163	8.9	94	0.00
36 M	1,2-Dichloroethane	0.433	0.407	6.0	99	0.00
37 M	Trichloroethene	0.391	0.366	6.4	104	0.00
38	Methylcyclohexane	0.495	0.461	6.9	106	0.00
39 M	1,2-Dichloropropane	0.341	0.315	7.6	100	0.00
40	Dibromomethane	0.228	0.216	5.3	102	0.00
41 M	Bromodichloromethane	0.442	0.403	8.8	101	0.00
42 M	Vinyl Acetate	0.745	0.695	6.7	101	0.00
43	Ethyl Acetate	0.367	0.333	9.3	100	0.00
44	Isopropyl Acetate	0.614	0.557	9.3	98	0.00
45 T	1,4-Dioxane	0.004	0.005#	-25.0	113	0.00
46	Methyl methacrylate	0.303	0.268	11.6	98	0.00
47	n-amyl Acetate	0.531	0.447	15.8	98	0.00
48 M	t-1,3-Dichloropropene	0.490	0.446	9.0	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011421\
 Data File : VN065501.D
 Acq On : 14 Jan 2021 11:10
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Jan 14 16:01:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jan 14 04:12:08 2021
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.539	0.497	7.8	103	0.00
50 M	1,1,2-Trichloroethane	0.334	0.318	4.8	104	0.00
51	Ethyl methacrylate	0.468	0.417	10.9	101	0.00
52	1,3-Dichloropropane	0.553	0.517	6.5	103	0.00
53 M	Dibromochloromethane	0.370	0.337	8.9	105	0.00
54 M	1,2-Dibromoethane	0.357	0.331	7.3	103	0.00
55 M	2-Chloroethyl vinyl ether	0.205	0.197	3.9	107	0.00
56 M	Bromoform	0.288	0.249	13.5	102	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
58 M	4-Methyl-2-Pentanone	0.398	0.366	8.0	97	0.00
59 M	2-Hexanone	0.288	0.264	8.3	99	0.00
60 S	4-Bromofluorobenzene	0.454	0.442	2.6	100	0.00
61 M	Tetrachloroethene	0.479	0.468	2.3	105	0.00
62 M	Toluene	1.566	1.455	7.1	102	0.00
63 S	Toluene-d8	1.313	1.322	-0.7	99	0.00
64 M	Chlorobenzene	1.018	0.947	7.0	103	0.00
65	1,1,1,2-Tetrachloroethane	0.383	0.352	8.1	101	0.00
66 M	Ethyl Benzene	1.712	1.558	9.0	102	0.00
67 M	m/p-Xylenes	0.670	0.616	8.1	103	0.00
68 M	o-Xylene	0.647	0.589	9.0	102	0.00
69 M	Styrene	1.081	0.985	8.9	103	0.00
70	Isopropylbenzene	1.692	1.536	9.2	102	0.00
71 M	1,1,2,2-Tetrachloroethane	0.481	0.428	11.0	98	0.00
72	1,2,3-Trichloropropane	0.457	0.409	10.5	100	0.00
73	Bromobenzene	0.484	0.435	10.1	101	0.00
74	n-propylbenzene	1.896	1.722	9.2	102	0.00
75	2-Chlorotoluene	1.144	1.029	10.1	100	0.00
76	1,3,5-Trimethylbenzene	1.450	1.332	8.1	103	0.00
77	t-1,4-Dichloro-2-butene	0.161	0.135	16.1	102	0.00
78	4-Chlorotoluene	1.157	1.037	10.4	101	0.00
79	tert-butylbenzene	1.246	1.118	10.3	101	0.00
80	1,2,4-Trimethylbenzene	1.443	1.315	8.9	102	0.00
81	sec-Butylbenzene	1.611	1.457	9.6	102	0.00
82	p-Isopropyltoluene	1.505	1.376	8.6	105	0.00
83 M	1,3-Dichlorobenzene	0.802	0.708	11.7	102	0.00
84 M	1,4-Dichlorobenzene	0.783	0.685	12.5	104	0.00
85	n-Butylbenzene	1.190	1.049	11.8	105	0.00
86 T	Hexachloroethane	0.248	0.206	16.9	99	0.00
87 M	1,2-Dichlorobenzene	0.790	0.714	9.6	103	0.00
88	1,2-Dibromo-3-Chloropropane	0.086	0.073	15.1	101	0.00
89	1,2,4-Trichlorobenzene	0.407	0.334	17.9	112	0.00
90	Hexachlorobutadiene	0.276	0.257	6.9	105	0.00
91 M	Naphthalene	1.005	0.784	22.0	112	0.00
92	1,2,3-Trichlorobenzene	0.401	0.333	17.0	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0