

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011421\
 Data File : VN065515.D
 Acq On : 14 Jan 2021 17:54
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Jan 15 04:15:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jan 14 04:12:08 2021
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	30.000	30.000	0.0	105	0.00
2 M	Dichlorodifluoromethane	20.000	17.012	14.9	98	0.00
3 M	Chloromethane	20.000	17.779	11.1	102	0.00
4 M	Vinyl Chloride	20.000	17.203	14.0	101	0.00
5 M	Bromomethane	20.000	16.414	17.9	92	0.00
6 M	Chloroethane	20.000	17.700	11.5	103	0.00
7 M	Trichlorofluoromethane	20.000	17.981	10.1	103	0.00
8 T	Diethyl Ether	20.000	18.186	9.1	102	0.00
9	1,1,2-Trichlorotrifluoroeth	20.000	17.934	10.3	101	0.00
10 M	1,1-Dichloroethene	20.000	18.012	9.9	104	0.00
11	Methyl Iodide	20.000	14.733	26.3	85	0.00
12	Methyl Acetate	20.000	17.816	10.9	99	0.00
13 M	Acrolein	100.000	87.699	12.3	102	0.00
14 M	Acrylonitrile	100.000	89.802	10.2	101	0.00
15 M	Acetone	100.000	92.822	7.2	102	0.00
16 M	Carbon Disulfide	20.000	16.942	15.3	101	0.00
17	Allyl chloride	20.000	17.344	13.3	102	0.00
18 M	Methylene Chloride	20.000	18.198	9.0	103	0.00
19 M	trans-1,2-Dichloroethene	20.000	17.799	11.0	103	0.00
20 T	Diisopropyl ether	20.000	18.078	9.6	103	0.00
21 M	1,1-Dichloroethane	20.000	18.197	9.0	104	0.00
22 M	cis-1,2-Dichloroethene	20.000	17.954	10.2	105	0.00
23 M	tert-Butyl Alcohol	100.000	82.994	17.0	92	0.00
24 M	Methyl tert-Butyl Ether	20.000	17.862	10.7	104	0.00
25 M	Chloroform	20.000	17.955	10.2	103	0.00
26	Cyclohexane	20.000	17.262	13.7	102	0.00
27 s	1,2-Dichloroethane-d4	30.000	29.650	1.2	101	0.00
28 I	1,4-Difluorobenzene	30.000	30.000	0.0	100	0.00
29	1,1-Dichloropropene	20.000	18.210	8.9	104	0.00
30 M	2-Butanone	100.000	89.697	10.3	96	0.00
31	2,2-Dichloropropane	20.000	18.138	9.3	102	0.00
32 M	1,1,1-Trichloroethane	20.000	18.415	7.9	102	0.00
33 M	Carbon Tetrachloride	20.000	18.251	8.7	104	0.00
34 M	Benzene	20.000	18.647	6.8	104	0.00
35	Methacrylonitrile	20.000	19.130	4.4	100	0.00
36 M	1,2-Dichloroethane	20.000	19.122	4.4	103	0.00
37 M	Trichloroethene	20.000	18.248	8.8	103	0.00
38	Methylcyclohexane	20.000	17.386	13.1	101	0.00
39 M	1,2-Dichloropropane	20.000	18.452	7.7	102	0.00
40	Dibromomethane	20.000	18.327	8.4	101	0.00
41 M	Bromodichloromethane	20.000	18.061	9.7	102	0.00
42 M	Vinyl Acetate	100.000	91.161	8.8	101	0.00
43	Ethyl Acetate	20.000	18.514	7.4	105	0.00
44	Isopropyl Acetate	20.000	18.059	9.7	100	0.00
45 T	1,4-Dioxane	400.000	352.982	11.8	97	0.00
46	Methyl methacrylate	20.000	17.727	11.4	100	0.00
47	n-amyl Acetate	20.000	17.200	14.0	102	0.00
48 M	t-1,3-Dichloropropene	20.000	17.197	14.0	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011421\
 Data File : VN065515.D
 Acq On : 14 Jan 2021 17:54
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Jan 15 04:15:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jan 14 04:12:08 2021
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	20.000	17.781	11.1	102	0.00
50 M	1,1,2-Trichloroethane	20.000	18.614	6.9	103	0.00
51	Ethyl methacrylate	20.000	17.791	11.0	102	0.00
52	1,3-Dichloropropane	20.000	18.576	7.1	104	0.00
53 M	Dibromochloromethane	20.000	17.461	12.7	102	0.00
54 M	1,2-Dibromoethane	20.000	17.850	10.7	102	0.00
55 M	2-Chloroethyl vinyl ether	100.000	93.642	6.4	107	0.00
56 M	Bromoform	20.000	16.659	16.7	101	0.00
57 I	Chlorobenzene-d5	30.000	30.000	0.0	102	0.00
58 M	4-Methyl-2-Pentanone	100.000	90.048	10.0	98	0.00
59 M	2-Hexanone	100.000	87.331	12.7	97	0.00
60 S	4-Bromofluorobenzene	30.000	29.069	3.1	102	0.00
61 M	Tetrachloroethene	20.000	18.448	7.8	102	0.00
62 M	Toluene	20.000	18.341	8.3	104	0.00
63 S	Toluene-d8	30.000	30.159	-0.5	102	0.00
64 M	Chlorobenzene	20.000	18.109	9.5	103	0.00
65	1,1,1,2-Tetrachloroethane	20.000	17.739	11.3	101	0.00
66 M	Ethyl Benzene	20.000	17.962	10.2	104	0.00
67 M	m/p-Xylenes	40.000	35.529	11.2	102	0.00
68 M	o-Xylene	20.000	18.142	9.3	105	0.00
69 M	Styrene	20.000	17.750	11.3	104	0.00
70	Isopropylbenzene	20.000	18.063	9.7	104	0.00
71 M	1,1,2,2-Tetrachloroethane	20.000	17.877	10.6	102	0.00
72	1,2,3-Trichloropropane	20.000	24.691	-23.5	141	0.00
73	Bromobenzene	20.000	17.760	11.2	103	0.00
74	n-propylbenzene	20.000	17.817	10.9	103	0.00
75	2-Chlorotoluene	20.000	17.808	11.0	102	0.00
76	1,3,5-Trimethylbenzene	20.000	17.935	10.3	103	0.00
77	t-1,4-Dichloro-2-butene	20.000	15.502	22.5	96	0.00
78	4-Chlorotoluene	20.000	17.605	12.0	102	0.00
79	tert-butylbenzene	20.000	17.788	11.1	103	0.00
80	1,2,4-Trimethylbenzene	20.000	17.727	11.4	102	0.00
81	sec-Butylbenzene	20.000	17.637	11.8	102	0.00
82	p-Isopropyltoluene	20.000	17.629	11.9	104	0.00
83 M	1,3-Dichlorobenzene	20.000	17.210	13.9	102	0.00
84 M	1,4-Dichlorobenzene	20.000	16.745	16.3	103	0.00
85	n-Butylbenzene	20.000	16.934	15.3	103	0.00
86 T	Hexachloroethane	20.000	16.533	17.3	101	0.00
87 M	1,2-Dichlorobenzene	20.000	17.747	11.3	104	0.00
88	1,2-Dibromo-3-Chloropropane	20.000	16.363	18.2	100	0.00
89	1,2,4-Trichlorobenzene	20.000	18.321	8.4	105	0.00
90	Hexachlorobutadiene	20.000	17.435	12.8	101	0.00
91 M	Naphthalene	20.000	18.712	6.4	106	0.00
92	1,2,3-Trichlorobenzene	20.000	15.508	22.5	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0