



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011421\
 Data File : VN065500.D
 Acq On : 14 Jan 2021 9:24
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011321W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Thu Jan 14 04:12:08 2021



AutoFind: Scans 3307, 3308, 3309; Background Corrected with Scan 3295

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.4	4012	PASS
75	95	30	60	45.1	11018	PASS
95	95	100	100	100.0	24443	PASS
96	95	5	9	6.6	1615	PASS
173	174	0.00	2	1.1	268	PASS
174	95	50	100	96.6	23613	PASS
175	174	5	9	7.1	1682	PASS
176	174	95	101	95.1	22459	PASS
177	176	5	9	6.6	1472	PASS