

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011422\
 Data File : VN070615.D
 Acq On : 14 Jan 2022 11:32
 Operator : JC/MD
 Sample : VN0114MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0114MBL01

Quant Time: Jan 14 13:35:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	814821	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1585182	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1889105	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	767253	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	600993	51.299	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.600%
35) Dibromofluoromethane	8.021	113	462536	48.225	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.460%
50) Toluene-d8	10.440	98	2109704	53.845	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.700%
62) 4-Bromofluorobenzene	12.731	95	924938	65.361	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	130.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011422\
Data File : VN070615.D
Acq On : 14 Jan 2022 11:32
Operator : JC/MD
Sample : VN0114MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0114MBL01

Quant Time: Jan 14 13:35:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

