

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011422\
 Data File : VN070620.D
 Acq On : 14 Jan 2022 13:47
 Operator : JC/MD
 Sample : N1119-01ME
 Misc : 5.57g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NORTH-EAST-CORNER-EXCAVATION-NORTH-SIDEWALLME

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 01/25/2022
 Supervised By :Mahesh Dadoda 01/25/2022

Quant Time: Jan 14 14:19:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	776990	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1317183	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1093451	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	508209	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	682162	61.062	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 122.120%			
35) Dibromofluoromethane	8.016	113	459063	57.602	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 115.200%			
50) Toluene-d8	10.440	98	1877940	57.682	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 115.360%			
62) 4-Bromofluorobenzene	12.731	95	738947	62.842	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 125.680%			
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.859	43	64680	3.083	ug/l #	71
31) Cyclohexane	8.083	56	210777m	11.956	ug/l	
39) Methylcyclohexane	9.453	83	270654	16.957	ug/l #	81
43) Isopropyl Acetate	8.550	43	82305	3.055	ug/l #	56
52) Toluene	10.502	92	31236	1.277	ug/l	95
67) Ethyl Benzene	11.846	91	1002828	23.591	ug/l	95
68) m/p-Xylenes	11.956	106	83628	5.405	ug/l	97
73) Isopropylbenzene	12.578	105	566780	12.009	ug/l	97
78) n-propylbenzene	12.921	91	1306670	24.989	ug/l	95
84) 1,2,4-Trimethylbenzene	13.366	105	94425	2.593	ug/l	95
85) sec-Butylbenzene	13.500	105	192110	4.249	ug/l #	66
95) Naphthalene	15.507	128	262248	12.715	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011422\
Data File : VN070620.D
Acq On : 14 Jan 2022 13:47
Operator : JC/MD
Sample : N1119-01ME
Misc : 5.57g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NORTH-EAST-CORNER-EXCAVATION-NORTH-SIDEWALLME

Quant Time: Jan 14 14:19:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 01/25/2022
Supervised By : Mahesh Dadoda 01/25/2022

