

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011422\
 Data File : VN070627.D
 Acq On : 14 Jan 2022 16:55
 Operator : JC/MD
 Sample : N1141-03MEDL 10X
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41862MEDL

Quant Time: Jan 17 05:40:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	792470	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1327136	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1240071	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	471986	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	654789	57.467	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 114.940%			
35) Dibromofluoromethane	8.021	113	452961	56.410	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 112.820%			
50) Toluene-d8	10.440	98	1933290	58.937	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 117.880%			
62) 4-Bromofluorobenzene	12.731	95	764611	64.537	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 129.080%			
Target Compounds					Qvalue	
16) Acetone	4.288	43	986164	133.326	ug/l	99
25) 2-Butanone	7.345	43	27692	2.973	ug/l	95
30) Chloroform	7.815	83	129882	7.285	ug/l	97
31) Cyclohexane	8.075	56	412705	22.953	ug/l #	62
39) Methylcyclohexane	9.459	83	406769	25.294	ug/l #	84
40) Benzene	8.458	78	602569	14.796	ug/l	100
52) Toluene	10.505	92	6699356	271.779	ug/l	95
67) Ethyl Benzene	11.846	91	1526024	31.654	ug/l	96
68) m/p-Xylenes	11.950	106	1932639	110.132	ug/l	92
69) o-Xylene	12.280	106	520767	29.836	ug/l	92
73) Isopropylbenzene	12.578	105	39026	0.890	ug/l	97
78) n-propylbenzene	12.921	91	69508	1.431	ug/l	96
80) 1,3,5-Trimethylbenzene	13.061	105	42900	1.221	ug/l	95
84) 1,2,4-Trimethylbenzene	13.369	105	111828	3.306	ug/l	100
95) Naphthalene	15.509	128	10659	2.275	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011422\
Data File : VN070627.D
Acq On : 14 Jan 2022 16:55
Operator : JC/MD
Sample : N1141-03MEDL 10X
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41862MEDL

Quant Time: Jan 17 05:40:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

