

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085441.D
 Acq On : 14 Jan 2025 16:07
 Operator : JC\MD
 Sample : VSTDIC010
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jan 15 01:44:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	195889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	340403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	294508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	132934	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	29856	9.442	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	18.880%#
35) Dibromofluoromethane	8.165	113	21110	8.939	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	17.880%#
50) Toluene-d8	10.565	98	73223	8.727	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	17.460%#
62) 4-Bromofluorobenzene	12.847	95	24327	8.476	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	16.960%#

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.124	85	26124	9.850	ug/l	96
3) Chloromethane	2.359	50	27143	9.452	ug/l	99
4) Vinyl Chloride	2.512	62	27862	9.653	ug/l	100
5) Bromomethane	2.965	94	17139	9.830	ug/l	87
6) Chloroethane	3.130	64	16791	9.175	ug/l	96
7) Trichlorofluoromethane	3.501	101	40755	9.731	ug/l	98
8) Diethyl Ether	3.953	74	12415	8.581	ug/l	94
9) 1,1,2-Trichlorotrifluo...	4.371	101	22982	9.742	ug/l	98
10) Methyl Iodide	4.589	142	25085	9.286	ug/l	99
11) Tert butyl alcohol	5.512	59	17962	49.608	ug/l	97
12) 1,1-Dichloroethene	4.342	96	20615	9.806	ug/l	99
13) Acrolein	4.177	56	25568	51.729	ug/l	99
14) Allyl chloride	5.018	41	32240	9.451	ug/l	97
15) Acrylonitrile	5.712	53	53959	46.918	ug/l	98
16) Acetone	4.424	43	48475	47.446	ug/l	98
17) Carbon Disulfide	4.712	76	60201	9.300	ug/l #	94
18) Methyl Acetate	5.030	43	30500	9.817	ug/l	98
19) Methyl tert-butyl Ether	5.789	73	65207	9.552	ug/l	91
20) Methylene Chloride	5.277	84	23723	9.379	ug/l	96
21) trans-1,2-Dichloroethene	5.783	96	21127	9.403	ug/l	97
22) Diisopropyl ether	6.671	45	71857	9.490	ug/l	97
23) Vinyl Acetate	6.600	43	250575	47.125	ug/l	98
24) 1,1-Dichloroethane	6.565	63	43099	9.335	ug/l	95
25) 2-Butanone	7.477	43	71191	47.350	ug/l	99
26) 2,2-Dichloropropane	7.483	77	36105	9.672	ug/l	99
27) cis-1,2-Dichloroethene	7.489	96	25049	9.466	ug/l	97
28) Bromochloromethane	7.806	49	19037	8.863	ug/l	97
29) Tetrahydrofuran	7.836	42	45641	47.881	ug/l	98
30) Chloroform	7.965	83	45792	9.596	ug/l	100
31) Cyclohexane	8.253	56	40211	10.020	ug/l #	93
32) 1,1,1-Trichloroethane	8.165	97	39170	9.358	ug/l	99
36) 1,1-Dichloropropene	8.371	75	30015	9.056	ug/l	97
37) Ethyl Acetate	7.559	43	30760	9.195	ug/l	97
38) Carbon Tetrachloride	8.359	117	36015	9.492	ug/l	94
39) Methylcyclohexane	9.594	83	27721	8.900	ug/l	95
40) Benzene	8.600	78	93680	9.407	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085441.D
 Acq On : 14 Jan 2025 16:07
 Operator : JC\MD
 Sample : VSTDIC010
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jan 15 01:44:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	16859	9.686	ug/l	96
42) 1,2-Dichloroethane	8.665	62	35569	9.483	ug/l	99
43) Isopropyl Acetate	8.683	43	48926	9.127	ug/l	99
44) Trichloroethene	9.347	130	21113	9.108	ug/l	95
45) 1,2-Dichloropropane	9.618	63	22747	8.942	ug/l	97
46) Dibromomethane	9.706	93	17260	9.404	ug/l	99
47) Bromodichloromethane	9.883	83	35020	9.365	ug/l	98
48) Methyl methacrylate	9.677	41	21873	9.067	ug/l	97
49) 1,4-Dioxane	9.694	88	7734	190.382	ug/l	98
51) 4-Methyl-2-Pentanone	10.441	43	146990	47.254	ug/l	98
52) Toluene	10.630	92	54976	9.527	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	32780	9.276	ug/l	96
54) cis-1,3-Dichloropropene	10.312	75	35904	9.512	ug/l	98
55) 1,1,2-Trichloroethane	11.012	97	21060	9.222	ug/l	98
56) Ethyl methacrylate	10.871	69	28454	9.042	ug/l	95
57) 1,3-Dichloropropane	11.159	76	37768	9.512	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	63110	43.550	ug/l	98
59) 2-Hexanone	11.194	43	101318	46.291	ug/l	98
60) Dibromochloromethane	11.353	129	25029	9.079	ug/l	98
61) 1,2-Dibromoethane	11.465	107	21055	9.263	ug/l	100
64) Tetrachloroethene	11.100	164	19934	9.928	ug/l	97
65) Chlorobenzene	11.888	112	61680	9.560	ug/l	93
66) 1,1,1,2-Tetrachloroethane	11.959	131	21853	9.229	ug/l	98
67) Ethyl Benzene	11.959	91	99248	9.445	ug/l	97
68) m/p-Xylenes	12.065	106	72431	18.650	ug/l	99
69) o-Xylene	12.394	106	34382	9.264	ug/l	99
70) Styrene	12.406	104	56285	9.163	ug/l	98
71) Bromoform	12.576	173	16071	9.485	ug/l #	98
73) Isopropylbenzene	12.694	105	87003	9.697	ug/l	99
74) N-amyl acetate	12.494	43	37343	9.260	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.935	83	30770	9.709	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	39589	14.675	ug/l	43
77) Bromobenzene	12.976	156	22784	9.720	ug/l	97
78) n-propylbenzene	13.035	91	102262	9.628	ug/l	98
79) 2-Chlorotoluene	13.123	91	68346	9.943	ug/l	100
80) 1,3,5-Trimethylbenzene	13.171	105	71656	9.669	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	9236	9.248	ug/l	100
82) 4-Chlorotoluene	13.218	91	66773	9.756	ug/l	97
83) tert-Butylbenzene	13.435	119	59455	9.553	ug/l	98
84) 1,2,4-Trimethylbenzene	13.476	105	72046	9.754	ug/l	99
85) sec-Butylbenzene	13.612	105	83662	9.699	ug/l	99
86) p-Isopropyltoluene	13.729	119	67556	9.674	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	41843	9.693	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	42728	9.549	ug/l	99
89) n-Butylbenzene	14.053	91	57632	9.314	ug/l	97
90) Hexachloroethane	14.329	117	15261	9.294	ug/l	97
91) 1,2-Dichlorobenzene	14.106	146	40726	9.458	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	5639	9.744	ug/l	95
93) 1,2,4-Trichlorobenzene	15.388	180	18720	9.354	ug/l	99
94) Hexachlorobutadiene	15.500	225	10491	9.874	ug/l	98
95) Naphthalene	15.635	128	52054	8.717	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	19450	9.605	ug/l	96


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085441.D
Acq On : 14 Jan 2025 16:07
Operator : JC\MD
Sample : VSTDIC010
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Jan 15 01:44:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:168 Resp: 195889
Ion Ratio Lower Upper
168 100
99 61.7 53.6 80.4



#2
Dichlorodifluoromethane
Concen: 9.850 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 85 Resp: 26124
Ion Ratio Lower Upper
85 100
87 35.8 16.7 50.1









#3
Chloromethane
Concen: 9.452 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 50 Resp: 27143
Ion Ratio Lower Upper
50 100
52 34.3 27.0 40.6



#4
Vinyl Chloride
Concen: 9.653 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 62 Resp: 27862
Ion Ratio Lower Upper
62 100
64 31.2 24.8 37.2









#5
Bromomethane
Concen: 9.830 ug/l
RT: 2.965 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 17139
Ion Ratio Lower Upper
94 100
96 102.1 71.8 107.6



#6
Chloroethane
Concen: 9.175 ug/l
RT: 3.130 min Scan# 200
Delta R.T. 0.012 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 64 Resp: 16791
Ion Ratio Lower Upper
64 100
66 33.1 24.6 36.8








#7
 Trichlorofluoromethane
 Concen: 9.731 ug/l
 RT: 3.501 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:101 Resp: 40755
 Ion Ratio Lower Upper
 101 100
 103 65.9 51.4 77.2



#8
 Diethyl Ether
 Concen: 8.581 ug/l
 RT: 3.953 min Scan# 340
 Delta R.T. 0.000 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07

Tgt Ion: 74 Resp: 12415
 Ion Ratio Lower Upper
 74 100
 45 105.3 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 9.742 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion:101 Resp: 22982
Ion Ratio Lower Upper
101 100
85 45.8 37.8 56.8
151 75.2 58.8 88.2



#10
Methyl Iodide
Concen: 9.286 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:142 Resp: 25085
Ion Ratio Lower Upper
142 100
127 50.9 40.3 60.5
141 15.0 12.8 19.2









#11
Tert butyl alcohol
Concen: 49.608 ug/l
RT: 5.512 min Scan# 605
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 59 Resp: 17962
Ion Ratio Lower Upper
59 100
57 11.9 8.6 13.0



#12
1,1-Dichloroethene
Concen: 9.806 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 96 Resp: 20615
Ion Ratio Lower Upper
96 100
61 176.8 141.0 211.4
98 60.5 49.0 73.6





#13

Acrolein

Concen: 51.729 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 56 Resp: 25568

Ion Ratio Lower Upper

56 100

55 70.9 56.3 84.5



#14

Allyl chloride

Concen: 9.451 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 41 Resp: 32240

Ion Ratio Lower Upper

41 100

39 78.3 64.4 96.6


76 35.1 30.5 45.7







#15
Acrylonitrile
Concen: 46.918 ug/l
RT: 5.712 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 53 Resp: 53959
Ion Ratio Lower Upper
53 100
52 83.3 65.5 98.3
51 38.8 29.8 44.8



#16
Acetone
Concen: 47.446 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07


Tgt Ion: 43 Resp: 48475
Ion Ratio Lower Upper
43 100
58 30.8 23.8 35.6








#17
Carbon Disulfide
Concen: 9.300 ug/l
RT: 4.712 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 76 Resp: 60201
Ion Ratio Lower Upper
76 100
78 10.7 6.9 10.3#



#18
Methyl Acetate
Concen: 9.817 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 43 Resp: 30500
Ion Ratio Lower Upper
43 100
74 24.3 18.6 28.0









#19
Methyl tert-butyl Ether
Concen: 9.552 ug/l
RT: 5.789 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 73 Resp: 65207
Ion Ratio Lower Upper
73 100
57 18.1 17.8 26.8



#20
Methylene Chloride
Concen: 9.379 ug/l
RT: 5.277 min Scan# 565
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 84 Resp: 23723
Ion Ratio Lower Upper
84 100
49 132.4 104.1 156.1
51 34.4 31.0 46.4
86 70.0 51.9 77.9









#21
trans-1,2-Dichloroethene
Concen: 9.403 ug/l
RT: 5.783 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 96 Resp: 21127
Ion Ratio Lower Upper
96 100
61 140.8 115.7 173.5
98 60.5 49.8 74.8



#22
Diisopropyl ether
Concen: 9.490 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 45 Resp: 71857
Ion Ratio Lower Upper
45 100
43 53.2 45.0 67.6
87 28.8 21.8 32.8
59 13.2 10.2 15.2









#23
 Vinyl Acetate
 Concen: 47.125 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.006 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 43 Resp: 250575
 Ion Ratio Lower Upper
 43 100
 86 9.8 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 9.335 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. 0.000 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07

Tgt Ion: 63 Resp: 43099
 Ion Ratio Lower Upper
 63 100
 98 8.8 3.3 9.8
 100 5.0 2.0 6.0









#25
2-Butanone
Concen: 47.350 ug/l
RT: 7.477 min Scan# 911
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 71191
Ion Ratio Lower Upper
43 100
72 24.8 20.2 30.4



#26
2,2-Dichloropropane
Concen: 9.672 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 77 Resp: 36105
Ion Ratio Lower Upper
77 100
97 20.8 10.7 32.1








#27
 cis-1,2-Dichloroethene
 Concen: 9.466 ug/l
 RT: 7.489 min Scan# 941
 Delta R.T. 0.006 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 96 Resp: 25049
 Ion Ratio Lower Upper
 96 100
 61 151.8 0.0 311.8
 98 64.7 0.0 126.0



#28
 Bromochloromethane
 Concen: 8.863 ug/l
 RT: 7.806 min Scan# 995
 Delta R.T. 0.000 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07

Tgt Ion: 49 Resp: 19037
 Ion Ratio Lower Upper
 49 100
 129 0.3 0.0 4.4
 130 70.8 55.0 82.4









#29
Tetrahydrofuran
Concen: 47.881 ug/l
RT: 7.836 min Scan# 1000
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 42 Resp: 45641
Ion Ratio Lower Upper
42 100
72 41.6 34.2 51.2
71 39.2 32.5 48.7



#30
Chloroform
Concen: 9.596 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 83 Resp: 45792
Ion Ratio Lower Upper
83 100
85 64.9 51.8 77.6





#31

Cyclohexane

Concen: 10.020 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 56 Resp: 40211

Ion Ratio Lower Upper

56 100

69 34.2 22.2 33.4#

84 87.2 66.4 99.6



#32

1,1,1-Trichloroethane

Concen: 9.358 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 97 Resp: 39170


Ion Ratio Lower Upper

97 100

99 62.6 49.8 74.6

61 50.2 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 9.442 ug/l

RT: 8.571 min Scan# 1125

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 65 Resp: 29856

Ion Ratio Lower Upper

65 100

67 49.5 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion: 114 Resp: 340403


Ion Ratio Lower Upper

114 100

63 21.1 0.0 47.6

88 15.5 0.0 32.6





#35

Dibromofluoromethane

Concen: 8.939 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min


Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Instrument :

MSVOA_N

ClientSampleId :



Tgt Ion:	113	Resp:	21110
Ion Ratio	Lower	Upper	
113	100		
111	103.4	82.7	124.1
192	16.4	14.3	21.5



#36

1,1-Dichloropropene

Concen: 9.056 ug/l




RT: 8.371 min Scan# 1091

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Tgt Ion:	75	Resp:	30015
Ion Ratio	Lower	Upper	
75	100		
110	34.3	16.5	49.5
77	32.9	24.4	36.6









#37
Ethyl Acetate
Concen: 9.195 ug/l
RT: 7.559 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 30760
Ion Ratio Lower Upper
43 100
61 12.2 10.9 16.3
70 8.4 7.5 11.3



#38
Carbon Tetrachloride
Concen: 9.492 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 117 Resp: 36015
Ion Ratio Lower Upper
117 100
119 101.2 75.4 113.2
121 29.1 24.6 37.0





#39

Methylcyclohexane

Concen: 8.900 ug/l

RT: 9.594 min Scan# 1129

Delta R.T. -0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 83 Resp: 27721

Ion Ratio Lower Upper

83 100

55 83.8 62.6 94.0

98 45.8 37.7 56.5



#40

Benzene

Concen: 9.407 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07

Tgt Ion: 78 Resp: 93680

Ion Ratio Lower Upper

78 100


77 22.0 19.0 28.6







#41
Methacrylonitrile
Concen: 9.686 ug/l
RT: 7.777 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion:	41	Resp:	16859
Ion	Ratio	Lower	Upper
41	100		
39	57.3	46.0	69.0
67	66.2	57.4	86.2
52	30.4	25.5	38.3



#42
1,2-Dichloroethane
Concen: 9.483 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:	62	Resp:	35569
Ion	Ratio	Lower	Upper
62	100		
98	8.2	0.0	17.0









#43
Isopropyl Acetate
Concen: 9.127 ug/l
RT: 8.683 min Scan# 1144
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 48926
Ion Ratio Lower Upper
43 100
61 26.2 20.7 31.1
87 12.8 9.8 14.8



#44
Trichloroethene
Concen: 9.108 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:130 Resp: 21113
Ion Ratio Lower Upper
130 100
95 92.9 0.0 195.8









#45
1,2-Dichloropropane
Concen: 8.942 ug/l
RT: 9.618 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 63 Resp: 22747
Ion Ratio Lower Upper
63 100
65 30.4 25.6 38.4



#46
Dibromomethane
Concen: 9.404 ug/l
RT: 9.706 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 93 Resp: 17260
Ion Ratio Lower Upper
93 100
95 81.5 64.7 97.1
174 87.8 69.0 103.4









#47
Bromodichloromethane
Concen: 9.365 ug/l
RT: 9.883 min Scan# 11
Delta R.T. 0.001 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 35020
Ion Ratio Lower Upper
83 100
85 62.6 51.2 76.8
127 6.9 6.5 9.7



#48
Methyl methacrylate
Concen: 9.067 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 41 Resp: 21873
Ion Ratio Lower Upper
41 100
69 80.0 64.7 97.1
39 65.2 49.0 73.6





#49

1,4-Dioxane

Concen: 190.382 ug/l

RT: 9.694 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 88 Resp: 7734

Ion Ratio Lower Upper

88 100

43 35.2 26.6 39.8

58 73.3 59.5 89.3



#50

Toluene-d8

Concen: 8.727 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion: 98 Resp: 73223

Ion Ratio Lower Upper

98 100

100 64.5 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 47.254 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 146990
Ion Ratio Lower Upper
43 100
58 37.1 30.5 45.7



#52
Toluene
Concen: 9.527 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 92 Resp: 54976
Ion Ratio Lower Upper
92 100
91 173.7 139.2 208.8









#53
 t-1,3-Dichloropropene
 Concen: 9.276 ug/l
 RT: 10.835 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 75 Resp: 32780
 Ion Ratio Lower Upper
 75 100
 77 27.2 23.5 35.3



#54
 cis-1,3-Dichloropropene
 Concen: 9.512 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. 0.006 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07


Tgt Ion: 75 Resp: 35904
 Ion Ratio Lower Upper
 75 100
 77 30.0 25.0 37.4
 39 53.1 43.1 64.7








#55
1,1,2-Trichloroethane
Concen: 9.222 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 97	Resp: 21060
Ion Ratio	Lower Upper
97 100	
83 88.3	69.0 103.6
85 57.3	44.6 66.8
99 60.2	48.6 72.8



#56
Ethyl methacrylate
Concen: 9.042 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 69	Resp: 28454		
Ion Ratio	Lower	Upper	
69 100			
41 72.1	54.6	82.0	
39 43.4	32.4	48.6	









#57
1,3-Dichloropropane
Concen: 9.512 ug/l
RT: 11.159 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 37768
Ion Ratio Lower Upper
76 100
78 31.1 25.6 38.4



#58
2-Chloroethyl Vinyl ether
Concen: 43.550 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 63 Resp: 63110
Ion Ratio Lower Upper
63 100
106 26.1 21.6 32.4









#59
2-Hexanone
Concen: 46.291 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 101318
Ion Ratio Lower Upper
43 100
58 50.9 26.2 78.6



#60
Dibromochloromethane
Concen: 9.079 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 129 Resp: 25029
Ion Ratio Lower Upper
129 100
127 79.1 38.6 115.8









#61
1,2-Dibromoethane
Concen: 9.263 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 21055
Ion Ratio Lower Upper
107 100
109 94.6 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 8.476 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 95 Resp: 24327
Ion Ratio Lower Upper
95 100
174 73.9 0.0 145.0
176 72.1 0.0 142.4








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:117 Resp: 294508
Ion Ratio Lower Upper
117 100
82 58.7 48.6 72.8
119 32.2 26.6 39.8



#64
Tetrachloroethene
Concen: 9.928 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:164 Resp: 19934
Ion Ratio Lower Upper
164 100
166 132.2 103.4 155.2
129 98.3 79.2 118.8
131 90.5 77.1 115.7









#65
Chlorobenzene
Concen: 9.560 ug/l
RT: 11.888 min Scan# 111
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:112 Resp: 61680
Ion Ratio Lower Upper
112 100
114 27.6 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 9.229 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:131 Resp: 21853
Ion Ratio Lower Upper
131 100
133 98.5 47.4 142.3
119 66.0 33.1 99.5









#67
Ethyl Benzene
Concen: 9.445 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 99248
Ion Ratio Lower Upper
91 100
106 28.2 23.8 35.8



#68
m/p-Xylenes
Concen: 18.650 ug/l
RT: 12.065 min Scan# 1719
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:106 Resp: 72431
Ion Ratio Lower Upper
106 100
91 211.7 167.7 251.5





#69

o-Xylene

Concen: 9.264 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:106 Resp: 34382

Ion Ratio Lower Upper

106 100

91 219.5 110.4 331.2



#70

Styrene

Concen: 9.163 ug/l

RT: 12.406 min Scan# 1777

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion:104 Resp: 56285


Ion Ratio Lower Upper

104 100

78 53.8 42.5 63.7

103 56.8 43.8 65.8





#71

Bromoform

Concen: 9.485 ug/l

RT: 12.576 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:173 Resp: 16071

Ion Ratio Lower Upper

173 100

175 50.0 24.4 73.2

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion:152 Resp: 132934


Ion Ratio Lower Upper

152 100

115 62.3 31.1 93.3

150 159.8 0.0 343.6









#73
Isopropylbenzene
Concen: 9.697 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:105 Resp: 87003
Ion Ratio Lower Upper
105 100
120 26.1 12.8 38.3



#74
N-ethyl acetate
Concen: 9.260 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 43 Resp: 37343
Ion Ratio Lower Upper
43 100
70 41.4 34.0 51.0
55 28.5 21.4 32.2
61 22.0 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 9.709 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 30770
Ion Ratio Lower Upper
83 100
131 9.6 4.8 14.4
85 63.6 32.2 96.6



#76
1,2,3-Trichloropropane
Concen: 14.675 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 75 Resp: 39589
Ion Ratio Lower Upper
75 100
77 127.6 109.7 329.2









#77
Bromobenzene
Concen: 9.720 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:156 Resp: 22784
Ion Ratio Lower Upper
156 100
77 221.6 114.1 342.4
158 98.7 48.9 146.8



#78
n-propylbenzene
Concen: 9.628 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 91 Resp: 102262
Ion Ratio Lower Upper
91 100
120 20.9 10.9 32.6









#79
2-Chlorotoluene
Concen: 9.943 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 68346
Ion Ratio Lower Upper
91 100
126 31.5 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 9.669 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:105 Resp: 71656
Ion Ratio Lower Upper
105 100
120 46.8 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 9.248 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 9236
Ion Ratio Lower Upper
75 100
53 119.6 95.6 143.4
89 46.6 37.0 55.6



#82
4-Chlorotoluene
Concen: 9.756 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 91 Resp: 66773
Ion Ratio Lower Upper
91 100
126 33.6 15.9 47.7





#83

tert-Butylbenzene

Concen: 9.553 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:119 Resp: 59455

Ion Ratio Lower Upper

119 100

91 70.0 35.1 105.3

134 27.8 12.0 36.1



#84

1,2,4-Trimethylbenzene

Concen: 9.754 ug/l

RT: 13.476 min Scan# 1959

Delta R.T. 0.000 min

Lab File: VN085441.D




Acq: 14 Jan 2025 16:07


Tgt Ion:105 Resp: 72046

Ion Ratio Lower Upper

105 100

120 43.8 21.6 65.0





#85

sec-Butylbenzene

Concen: 9.699 ug/l

RT: 13.612 min Scan# 1982

Delta R.T. -0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:105 Resp: 83662

Ion Ratio Lower Upper

105 100

134 18.9 9.7 28.9



#86

p-Isopropyltoluene

Concen: 9.674 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.006 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion:119 Resp: 67556


Ion Ratio Lower Upper

119 100

134 25.2 12.7 38.0

91 25.0 12.7 38.1





#87

1,3-Dichlorobenzene

Concen: 9.693 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:146 Resp: 41843

Ion Ratio Lower Upper

146 100

111 43.1 21.4 64.3

148 63.0 32.3 96.9



#88

1,4-Dichlorobenzene

Concen: 9.549 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion:146 Resp: 42728


Ion Ratio Lower Upper

146 100

111 44.0 21.3 63.7

148 64.6 32.4 97.0





#89

n-Butylbenzene

Concen: 9.314 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN085441.D





Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:	91	Resp:	57632
Ion	Ratio	Lower	Upper
91	100		
92	49.2	25.8	77.3
134	22.8	11.7	35.1



#90

Hexachloroethane

Concen: 9.294 ug/l




RT: 14.329 min Scan# 2104


Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Tgt Ion:	117	Resp:	15261
Ion	Ratio	Lower	Upper
117	100		
201	65.0	33.7	101.0








#91
 1,2-Dichlorobenzene
 Concen: 9.458 ug/l
 RT: 14.106 min Scan# 2066
 Delta R.T. 0.006 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:146 Resp: 40726
 Ion Ratio Lower Upper
 146 100
 111 42.9 21.7 65.1
 148 62.9 31.4 94.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 9.744 ug/l
 RT: 14.717 min Scan# 2170
 Delta R.T. 0.000 min
 Lab File: VN085441.D
 Acq: 14 Jan 2025 16:07

Tgt Ion: 75 Resp: 5639
 Ion Ratio Lower Upper
 75 100
 155 67.9 36.4 109.2
 157 94.1 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 9.354 ug/l

RT: 15.388 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:180 Resp: 18720

Ion Ratio Lower Upper

180 100

182 92.4 46.8 140.3

145 31.7 16.0 48.0



#94

Hexachlorobutadiene

Concen: 9.874 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07




Tgt Ion:225 Resp: 10491


Ion Ratio Lower Upper

225 100

223 63.3 30.7 92.1

227 62.1 30.9 92.5





#95

Naphthalene

Concen: 8.717 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. 0.000 min

Lab File: VN085441.D





Acq: 14 Jan 2025 16:07

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:128	Resp:	52054
Ion Ratio	Lower	Upper
128	100	
127	12.6	10.6 16.0
129	11.2	8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 9.605 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. 0.000 min

Lab File: VN085441.D

Acq: 14 Jan 2025 16:07

Tgt Ion:180	Resp:	19450
Ion Ratio	Lower	Upper
180	100	
182	90.3	47.4 142.2
145	34.7	16.9 50.7

