

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085444.D
 Acq On : 14 Jan 2025 17:42
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 15 02:20:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153702	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	275988	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	234940	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	88711	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	136365	54.962	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.920%		
35) Dibromofluoromethane	8.165	113	101523	53.024	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.040%		
50) Toluene-d8	10.565	98	335165	49.269	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 98.540%		
62) 4-Bromofluorobenzene	12.847	95	98696	42.412	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 84.820%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085444.D
Acq On : 14 Jan 2025 17:42
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 15 02:20:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

