

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085438.D
 Acq On : 14 Jan 2025 14:56
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Jan 15 01:41:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	202613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	329363	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	298938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	146196	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	313645	95.899	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	191.800%#
35) Dibromofluoromethane	8.159	113	236159	103.353	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	206.700%#
50) Toluene-d8	10.565	98	881956	108.636	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	217.280%#
62) 4-Bromofluorobenzene	12.847	95	313161	112.765	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	225.520%#

Target Compounds					Qvalue
2) Dichlorodifluoromethane	2.124	85	269023	98.065	ug/l 99
3) Chloromethane	2.359	50	275415	92.724	ug/l 98
4) Vinyl Chloride	2.512	62	282361	94.576	ug/l 99
5) Bromomethane	2.942	94	158888	88.105	ug/l 98
6) Chloroethane	3.112	64	176111	93.043	ug/l 99
7) Trichlorofluoromethane	3.494	101	423967	97.870	ug/l 100
8) Diethyl Ether	3.953	74	146889	98.154	ug/l 98
9) 1,1,2-Trichlorotrifluo...	4.371	101	239104	97.997	ug/l 99
10) Methyl Iodide	4.589	142	292983	104.856	ug/l 95
11) Tert butyl alcohol	5.512	59	173044	462.061	ug/l 99
12) 1,1-Dichloroethene	4.336	96	222117	102.146	ug/l 99
13) Acrolein	4.171	56	275350	538.600	ug/l 99
14) Allyl chloride	5.018	41	349717	99.111	ug/l 98
15) Acrylonitrile	5.712	53	585370	492.095	ug/l 99
16) Acetone	4.418	43	483221	457.268	ug/l 96
17) Carbon Disulfide	4.712	76	630294	94.140	ug/l 100
18) Methyl Acetate	5.012	43	304226	94.674	ug/l 99
19) Methyl tert-butyl Ether	5.788	73	743008	105.230	ug/l 100
20) Methylene Chloride	5.271	84	254912	97.440	ug/l 98
21) trans-1,2-Dichloroethene	5.783	96	231503	99.618	ug/l 97
22) Diisopropyl ether	6.665	45	820871	104.812	ug/l 99
23) Vinyl Acetate	6.594	43	3029568	550.850	ug/l 99
24) 1,1-Dichloroethane	6.565	63	471831	98.803	ug/l 99
25) 2-Butanone	7.477	43	764979	491.912	ug/l 98
26) 2,2-Dichloropropane	7.482	77	388140	100.530	ug/l 100
27) cis-1,2-Dichloroethene	7.482	96	280197	102.367	ug/l 99
28) Bromochloromethane	7.806	49	214796	96.679	ug/l 100
29) Tetrahydrofuran	7.830	42	503459	510.635	ug/l 99
30) Chloroform	7.959	83	484941	98.253	ug/l 100
31) Cyclohexane	8.253	56	398641	96.041	ug/l 96
32) 1,1,1-Trichloroethane	8.165	97	426782	98.578	ug/l 98
36) 1,1-Dichloropropene	8.365	75	342961	106.945	ug/l 100
37) Ethyl Acetate	7.553	43	326802	100.961	ug/l 97
38) Carbon Tetrachloride	8.359	117	378390	103.065	ug/l 97
39) Methylcyclohexane	9.594	83	371734	123.354	ug/l 99
40) Benzene	8.600	78	1021393	106.001	ug/l 100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085438.D
 Acq On : 14 Jan 2025 14:56
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Jan 15 01:41:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	184326	109.454	ug/l	96
42) 1,2-Dichloroethane	8.665	62	374938	103.314	ug/l	100
43) Isopropyl Acetate	8.682	43	549969	106.035	ug/l	100
44) Trichloroethene	9.347	130	238296	106.242	ug/l	96
45) 1,2-Dichloropropane	9.618	63	256964	104.405	ug/l	97
46) Dibromomethane	9.706	93	180990	101.913	ug/l	99
47) Bromodichloromethane	9.882	83	388499	107.377	ug/l	98
48) Methyl methacrylate	9.676	41	267689	114.685	ug/l	99
49) 1,4-Dioxane	9.688	88	83879	2133.992	ug/l	97
51) 4-Methyl-2-Pentanone	10.441	43	1643800	546.159	ug/l	100
52) Toluene	10.623	92	634743	113.689	ug/l	99
53) t-1,3-Dichloropropene	10.829	75	391238	114.424	ug/l	97
54) cis-1,3-Dichloropropene	10.306	75	410403	112.369	ug/l	99
55) 1,1,2-Trichloroethane	11.012	97	229522	103.879	ug/l	98
56) Ethyl methacrylate	10.870	69	383663	101.029	ug/l	99
57) 1,3-Dichloropropane	11.159	76	412886	107.469	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.153	63	848636	605.243	ug/l	100
59) 2-Hexanone	11.194	43	1178831	556.643	ug/l	99
60) Dibromochloromethane	11.353	129	283361	106.227	ug/l	99
61) 1,2-Dibromoethane	11.465	107	230104	104.623	ug/l	98
64) Tetrachloroethene	11.100	164	209870	102.980	ug/l	98
65) Chlorobenzene	11.888	112	677679	103.482	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	241593	100.517	ug/l	99
67) Ethyl Benzene	11.959	91	1239080	116.169	ug/l	99
68) m/p-Xylenes	12.070	106	927244	235.219	ug/l	99
69) o-Xylene	12.394	106	441279	117.135	ug/l	100
70) Styrene	12.406	104	760136	121.919	ug/l	99
71) Bromoform	12.576	173	186124	108.227	ug/l #	99
73) Isopropylbenzene	12.694	105	1146752	116.221	ug/l	100
74) N-amyl acetate	12.488	43	489217	110.308	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	327767	94.039	ug/l	100
76) 1,2,3-Trichloropropane	12.988	75	310115m	104.530	ug/l	
77) Bromobenzene	12.976	156	268616	104.200	ug/l	97
78) n-propylbenzene	13.029	91	1385353	118.600	ug/l	100
79) 2-Chlorotoluene	13.123	91	822258	108.767	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	953066	116.938	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	122508	111.544	ug/l	87
82) 4-Chlorotoluene	13.217	91	832317	110.574	ug/l	99
83) tert-Butylbenzene	13.435	119	797069	116.457	ug/l	98
84) 1,2,4-Trimethylbenzene	13.476	105	971729	119.625	ug/l	100
85) sec-Butylbenzene	13.611	105	1143885	120.587	ug/l	100
86) p-Isopropyltoluene	13.723	119	960747	100.111	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	502960	105.943	ug/l	99
88) 1,4-Dichlorobenzene	13.811	146	498709	101.341	ug/l	97
89) n-Butylbenzene	14.053	91	842368	123.781	ug/l	98
90) Hexachloroethane	14.329	117	188541	104.405	ug/l	100
91) 1,2-Dichlorobenzene	14.100	146	470919	99.445	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	63846	100.313	ug/l	98
93) 1,2,4-Trichlorobenzene	15.388	180	250787	113.946	ug/l	99
94) Hexachlorobutadiene	15.500	225	112899	96.619	ug/l	99
95) Naphthalene	15.635	128	756816	115.234	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	238860	107.256	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085438.D
Acq On : 14 Jan 2025 14:56
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:41:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085438.D
Acq On : 14 Jan 2025 14:56
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Quant Time: Jan 15 01:41:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:168 Resp: 202613
Ion Ratio Lower Upper
168 100
99 64.9 53.6 80.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#2
Dichlorodifluoromethane
Concen: 98.065 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56


Tgt Ion: 85 Resp: 269023
Ion Ratio Lower Upper
85 100
87 32.7 16.7 50.1





#3
Chloromethane
Concen: 92.724 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 50 Resp: 275415
Ion Ratio Lower Upper
50 100
52 32.4 27.0 40.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#4
Vinyl Chloride
Concen: 94.576 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 62 Resp: 282361
Ion Ratio Lower Upper
62 100
64 31.6 24.8 37.2









#5
Bromomethane
Concen: 88.105 ug/l
RT: 2.942 min Scan# 16
Delta R.T. -0.012 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 94 Resp: 158888
Ion Ratio Lower Upper
94 100
96 91.9 71.8 107.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#6
Chloroethane
Concen: 93.043 ug/l
RT: 3.112 min Scan# 197
Delta R.T. -0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 64 Resp: 176111
Ion Ratio Lower Upper
64 100
66 31.1 24.6 36.8






#7
 Trichlorofluoromethane
 Concen: 97.870 ug/l
 RT: 3.494 min Scan# 263
 Delta R.T. -0.006 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100






Tgt Ion:101 Resp: 423967
 Ion Ratio Lower Upper
 101 100
 103 64.4 51.4 77.2

Manual Integrations




APPROVED


Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025



#8
 Diethyl Ether
 Concen: 98.154 ug/l
 RT: 3.953 min Scan# 340
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56


Tgt Ion: 74 Resp: 146889
 Ion Ratio Lower Upper
 74 100
 45 97.4 49.7 149.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 97.997 ug/l
RT: 4.371 min Scan# 411
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100




Tgt Ion:101 Resp: 239104
Ion Ratio Lower Upper
101 100
85 47.1 37.8 56.8
151 71.8 58.8 88.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#10
Methyl Iodide
Concen: 104.856 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56



Tgt Ion:142 Resp: 292983
Ion Ratio Lower Upper
142 100
127 47.1 40.3 60.5
141 13.1 12.8 19.2





#11

Tert butyl alcohol

Concen: 462.061 ug/l

RT: 5.512 min Scan# 605

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 59 Resp: 173044

Ion Ratio Lower Upper

59 100





57 11.1 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#12

1,1-Dichloroethene

Concen: 102.146 ug/l

RT: 4.336 min Scan# 405

Delta R.T. -0.006 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion: 96 Resp: 222117


Ion Ratio Lower Upper

96 100

61 175.3 141.0 211.4

98 62.7 49.0 73.6









#13
 Acrolein
 Concen: 538.600 ug/l
 RT: 4.171 min Scan# 377
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Tgt Ion: 56 Resp: 275350
 Ion Ratio Lower Upper
 56 100
 55 71.1 56.3 84.5




Manual Integrations
 APPROVED


Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025



#14
 Allyl chloride
 Concen: 99.111 ug/l
 RT: 5.018 min Scan# 521
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56

Tgt Ion: 41 Resp: 349717
 Ion Ratio Lower Upper
 41 100
 39 83.3 64.4 96.6
 76 37.9 30.5 45.7









#15
Acrylonitrile
Concen: 492.095 ug/l
RT: 5.712 min Scan# 639
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion	Ratio	Lower	Upper
53	100		
52	83.0	65.5	98.3
51	37.1	29.8	44.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#16
Acetone
Concen: 457.268 ug/l
RT: 4.418 min Scan# 419
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion	Ratio	Lower	Upper
43	100		
58	27.3	23.8	35.6









#17
Carbon Disulfide
Concen: 94.140 ug/l
RT: 4.712 min Scan# 468
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 76 Resp: 630294
Ion Ratio Lower Upper
76 100
78 8.7 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#18
Methyl Acetate
Concen: 94.674 ug/l
RT: 5.012 min Scan# 520
Delta R.T. -0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 43 Resp: 304226
Ion Ratio Lower Upper
43 100
74 23.6 18.6 28.0









#19
Methyl tert-butyl Ether
Concen: 105.230 ug/l
RT: 5.788 min Scan# 652
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 73 Resp: 743008
Ion Ratio Lower Upper
73 100
57 22.3 17.8 26.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#20
Methylene Chloride
Concen: 97.440 ug/l
RT: 5.271 min Scan# 564
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56


Tgt Ion: 84 Resp: 254912
Ion Ratio Lower Upper
84 100
49 128.5 104.1 156.1
51 37.8 31.0 46.4
86 61.1 51.9 77.9





#21
trans-1,2-Dichloroethene
Concen: 99.618 ug/l
RT: 5.783 min Scan# 651
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 96 Resp: 231503
Ion Ratio Lower Upper
96 100
61 148.3 115.7 173.5
98 65.0 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#22
Diisopropyl ether
Concen: 104.812 ug/l
RT: 6.665 min Scan# 801
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 45 Resp: 820871
Ion Ratio Lower Upper
45 100
43 55.7 45.0 67.6
87 27.2 21.8 32.8
59 11.6 10.2 15.2









#23
 Vinyl Acetate
 Concen: 550.850 ug/l
 RT: 6.594 min Scan# 789
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Tgt Ion: 43 Resp: 3029568
 Ion Ratio Lower Upper
 43 100
 86 10.1 8.4 12.6




Manual Integrations
 APPROVED


Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025



#24
 1,1-Dichloroethane
 Concen: 98.803 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56


Tgt Ion: 63 Resp: 471831
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.3 9.8
 100 4.6 2.0 6.0





#25
2-Butanone
Concen: 491.912 ug/l
RT: 7.477 min Scan# 939
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100




Tgt Ion: 43 Resp: 764979
Ion Ratio Lower Upper
43 100
72 26.3 20.2 30.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#26
2,2-Dichloropropane
Concen: 100.530 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Tgt Ion: 77 Resp: 388140
Ion Ratio Lower Upper
77 100
97 21.4 10.7 32.1





#27
cis-1,2-Dichloroethene
Concen: 102.367 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 96 Resp: 280197
Ion Ratio Lower Upper
96 100
61 156.8 0.0 311.8
98 65.0 0.0 126.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#28
Bromochloromethane
Concen: 96.679 ug/l
RT: 7.806 min Scan# 995
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 49 Resp: 214796
Ion Ratio Lower Upper
49 100
129 1.9 0.0 4.4
130 68.4 55.0 82.4









#29
Tetrahydrofuran
Concen: 510.635 ug/l
RT: 7.830 min Scan# 999
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 42 Resp: 503459
Ion Ratio Lower Upper
42 100
72 43.1 34.2 51.2
71 40.9 32.5 48.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#30
Chloroform
Concen: 98.253 ug/l
RT: 7.959 min Scan# 1021
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56


Tgt Ion: 83 Resp: 484941
Ion Ratio Lower Upper
83 100
85 65.0 51.8 77.6





#31
Cyclohexane
Concen: 96.041 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion	Ratio	Lower	Upper
56	100		
69	29.2	22.2	33.4
84	86.4	66.4	99.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#32
1,1,1-Trichloroethane
Concen: 98.578 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56


Tgt Ion	Ratio	Lower	Upper
97	100		
99	64.2	49.8	74.6
61	50.6	41.4	62.2





#33
1,2-Dichloroethane-d4
Concen: 95.899 ug/l
RT: 8.571 min Scan# 1125
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100




Tgt Ion: 65 Resp: 313645
Ion Ratio Lower Upper
65 100
67 51.1 0.0 101.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.094 min Scan# 1214
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56



Tgt Ion:114 Resp: 329363
Ion Ratio Lower Upper
114 100
63 21.2 0.0 47.6
88 16.2 0.0 32.6






#35
 Dibromofluoromethane
 Concen: 103.353 ug/l
 RT: 8.159 min Scan# 1055
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC100




Tgt Ion:	113	Resp:	236159
Ion	Ratio	Lower	Upper
113	100		
111	102.3	82.7	124.1
192	17.4	14.3	21.5

Manual Integrations
 APPROVED



Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025




#36
 1,1-Dichloropropene
 Concen: 106.945 ug/l
 RT: 8.365 min Scan# 1090
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56



Tgt Ion:	75	Resp:	342961
Ion	Ratio	Lower	Upper
75	100		
110	32.6	16.5	49.5
77	30.6	24.4	36.6









#37
Ethyl Acetate
Concen: 100.961 ug/l
RT: 7.553 min Scan# 952
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 326802
Ion Ratio Lower Upper
43 100
61 12.0 10.9 16.3
70 9.6 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#38
Carbon Tetrachloride
Concen: 103.065 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:117 Resp: 378390
Ion Ratio Lower Upper
117 100
119 98.1 75.4 113.2
121 30.2 24.6 37.0









#39
Methylcyclohexane
Concen: 123.354 ug/l
RT: 9.594 min Scan# 1299
Delta R.T. -0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 83 Resp: 371734
Ion Ratio Lower Upper
83 100
55 79.3 62.6 94.0
98 46.7 37.7 56.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#40
Benzene
Concen: 106.001 ug/l
RT: 8.600 min Scan# 1130
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 78 Resp: 1021393
Ion Ratio Lower Upper
78 100
77 23.9 19.0 28.6









#41
Methacrylonitrile
Concen: 109.454 ug/l
RT: 7.777 min Scan# 99
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:	41	Resp:	184326
Ion	Ratio	Lower	Upper
41	100		
39	52.8	46.0	69.0
67	68.8	57.4	86.2
52	32.3	25.5	38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#42
1,2-Dichloroethane
Concen: 103.314 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:	62	Resp:	374938
Ion	Ratio	Lower	Upper
62	100		
98	8.4	0.0	17.0





#43

Isopropyl Acetate

Concen: 106.035 ug/l

RT: 8.682 min Scan# 1144

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100





Tgt Ion:	43	Resp:	549969
Ion	Ratio	Lower	Upper
43	100		
61	25.7	20.7	31.1
87	12.2	9.8	14.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#44

Trichloroethene

Concen: 106.242 ug/l




RT: 9.347 min Scan# 1257


Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion:	130	Resp:	238296
Ion	Ratio	Lower	Upper
130	100		
95	101.5	0.0	195.8





#45

1,2-Dichloropropane

Concen: 104.405 ug/l

RT: 9.618 min Scan# 13

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 63 Resp: 256964

Ion Ratio Lower Upper

63 100





65 30.3 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#46

Dibromomethane

Concen: 101.913 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion: 93 Resp: 180990


Ion Ratio Lower Upper

93 100

95 81.6 64.7 97.1

174 87.5 69.0 103.4






#47
 Bromodichloromethane
 Concen: 107.377 ug/l
 RT: 9.882 min Scan# 1348
 Delta R.T. 0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :




VSTDICC100



Tgt Ion:	83	Resp:	388499
Ion	Ratio	Lower	Upper
83	100		
85	65.9	51.2	76.8
127	8.4	6.5	9.7




Manual Integrations
 APPROVED


Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025



#48
 Methyl methacrylate
 Concen: 114.685 ug/l
 RT: 9.676 min Scan# 1313
 Delta R.T. -0.001 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56


Tgt Ion:	41	Resp:	267689
Ion	Ratio	Lower	Upper
41	100		
69	80.9	64.7	97.1
39	63.1	49.0	73.6





#49
1,4-Dioxane
Concen: 2133.992 ug/l
RT: 9.688 min Scan# 1315
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion: 88 Resp: 83879
Ion Ratio Lower Upper
88 100
43 31.3 26.6 39.8
58 72.1 59.5 89.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#50
Toluene-d8
Concen: 108.636 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 98 Resp: 881956
Ion Ratio Lower Upper
98 100
100 64.5 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 546.159 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion: 43 Resp: 1643800
Ion Ratio Lower Upper
43 100
58 38.4 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#52
Toluene
Concen: 113.689 ug/l
RT: 10.623 min Scan# 1474
Delta R.T. -0.001 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 92 Resp: 634743
Ion Ratio Lower Upper
92 100
91 173.1 139.2 208.8









#53
 t-1,3-Dichloropropene
 Concen: 114.424 ug/l
 RT: 10.829 min Scan# 1510
 Delta R.T. -0.006 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100

Tgt Ion: 75 Resp: 391238
 Ion Ratio Lower Upper
 75 100
 77 31.0 23.5 35.3




Manual Integrations
 APPROVED


Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025



#54
 cis-1,3-Dichloropropene
 Concen: 112.369 ug/l
 RT: 10.306 min Scan# 1420
 Delta R.T. -0.000 min
 Lab File: VN085438.D
 Acq: 14 Jan 2025 14:56


Tgt Ion: 75 Resp: 410403
 Ion Ratio Lower Upper
 75 100
 77 31.9 25.0 37.4
 39 54.3 43.1 64.7





#55
1,1,2-Trichloroethane
Concen: 103.879 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100




Tgt Ion:	97	Resp:	229522
Ion	Ratio	Lower	Upper
97	100		
83	88.4	69.0	103.6
85	56.3	44.6	66.8
99	62.9	48.6	72.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#56
Ethyl methacrylate
Concen: 101.029 ug/l
RT: 10.870 min Scan# 1516
Delta R.T. -0.001 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Tgt Ion:	69	Resp:	383663
Ion	Ratio	Lower	Upper
69	100		
41	69.1	54.6	82.0
39	40.1	32.4	48.6





#57
1,3-Dichloropropane
Concen: 107.469 ug/l
RT: 11.159 min Scan# 1565
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100




Tgt Ion: 76 Resp: 412886
Ion Ratio Lower Upper
76 100
78 32.2 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#58
2-Chloroethyl Vinyl ether
Concen: 605.243 ug/l
RT: 10.153 min Scan# 1394
Delta R.T. -0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56



Tgt Ion: 63 Resp: 848636
Ion Ratio Lower Upper
63 100
106 26.8 21.6 32.4









#59
2-Hexanone
Concen: 556.643 ug/l
RT: 11.194 min Scan# 1570
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 1178831
Ion Ratio Lower Upper
43 100
58 53.2 26.2 78.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#60
Dibromochloromethane
Concen: 106.227 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56


Tgt Ion:129 Resp: 283361
Ion Ratio Lower Upper
129 100
127 78.1 38.6 115.8





#61
1,2-Dibromoethane
Concen: 104.623 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion: 107 Resp: 230104
Ion Ratio Lower Upper
107 100
109 93.1 75.9 113.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025



#62
4-Bromofluorobenzene
Concen: 112.765 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 95 Resp: 313161
Ion Ratio Lower Upper
95 100
174 73.4 0.0 145.0
176 70.5 0.0 142.4








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion:	117	Resp:	298938
Ion Ratio	Lower	Upper	
117	100		
82	59.3	48.6	72.8
119	31.6	26.6	39.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#64
Tetrachloroethene
Concen: 102.980 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:	164	Resp:	209870
Ion Ratio	Lower	Upper	
164	100		
166	126.8	103.4	155.2
129	98.3	79.2	118.8
131	94.7	77.1	115.7









#65
Chlorobenzene
Concen: 103.482 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion:112 Resp: 677679
Ion Ratio Lower Upper
112 100
114 32.5 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 100.517 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:131 Resp: 241593
Ion Ratio Lower Upper
131 100
133 94.3 47.4 142.3
119 64.9 33.1 99.5









#67
Ethyl Benzene
Concen: 116.169 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 91 Resp: 1239080
Ion Ratio Lower Upper
91 100
106 30.3 23.8 35.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#68
m/p-Xylenes
Concen: 235.219 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:106 Resp: 927244
Ion Ratio Lower Upper
106 100
91 208.0 167.7 251.5









#69
o-Xylene
Concen: 117.135 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:106 Resp: 441279
Ion Ratio Lower Upper
106 100
91 221.3 110.4 331.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#70
Styrene
Concen: 121.919 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:104 Resp: 760136
Ion Ratio Lower Upper
104 100
78 52.8 42.5 63.7
103 55.3 43.8 65.8





#71

Bromoform

Concen: 108.227 ug/l

RT: 12.576 min Scan# 18

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100





Tgt Ion:	173	Resp:	186124
Ion Ratio	Lower	Upper	
173	100		
175	48.4	24.4	73.2
254	0.1	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion:	152	Resp:	146196
Ion Ratio	Lower	Upper	
152	100		
115	62.7	31.1	93.3
150	193.1	0.0	343.6









#73
Isopropylbenzene
Concen: 116.221 ug/l
RT: 12.694 min Scan# 1825
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:105 Resp: 1146752
Ion Ratio Lower Upper
105 100
120 25.7 12.8 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#74
N-ethyl acetate
Concen: 110.308 ug/l
RT: 12.488 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 43 Resp: 489217
Ion Ratio Lower Upper
43 100
70 41.9 34.0 51.0
55 25.7 21.4 32.2
61 23.5 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 94.039 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100





Tgt Ion:	83	Resp:	327767
Ion	Ratio	Lower	Upper
83	100		
131	9.8	4.8	14.4
85	64.1	32.2	96.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#76

1,2,3-Trichloropropane

Concen: 104.530 ug/l m




RT: 12.988 min Scan# 1876


Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion:	75	Resp:	310115
Ion	Ratio	Lower	Upper
75	100		
77	192.6	109.7	329.2





#77

Bromobenzene

Concen: 104.200 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion:	156	Resp:	268616
Ion Ratio	Lower	Upper	
156	100		
77	222.4	114.1	342.4
158	96.3	48.9	146.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#78

n-propylbenzene

Concen: 118.600 ug/l




RT: 13.029 min Scan# 1883


Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Tgt Ion:	91	Resp:	1385353
Ion Ratio	Lower	Upper	
91	100		
120	21.8	10.9	32.6









#79
2-Chlorotoluene
Concen: 108.767 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 91 Resp: 822258
Ion Ratio Lower Upper
91 100
126 31.8 15.7 47.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#80
1,3,5-Trimethylbenzene
Concen: 116.938 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56


Tgt Ion:105 Resp: 953066
Ion Ratio Lower Upper
105 100
120 47.8 23.9 71.7





#81
trans-1,4-Dichloro-2-butene
Concen: 111.544 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 75 Resp: 122508
Ion Ratio Lower Upper
75 100
53 100.9 95.6 143.4
89 44.4 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#82
4-Chlorotoluene
Concen: 110.574 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 91 Resp: 832317
Ion Ratio Lower Upper
91 100
126 32.5 15.9 47.7









#83
tert-Butylbenzene
Concen: 116.457 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:119 Resp: 797069
Ion Ratio Lower Upper
119 100
91 70.7 35.1 105.3
134 26.6 12.0 36.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#84
1,2,4-Trimethylbenzene
Concen: 119.625 ug/l
RT: 13.476 min Scan# 1959
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:105 Resp: 971729
Ion Ratio Lower Upper
105 100
120 43.4 21.6 65.0









#85
sec-Butylbenzene
Concen: 120.587 ug/l
RT: 13.611 min Scan# 1982
Delta R.T. -0.001 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:105 Resp: 1143885
Ion Ratio Lower Upper
105 100
134 19.2 9.7 28.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#86
p-Isopropyltoluene
Concen: 100.111 ug/l
RT: 13.723 min Scan# 2001
Delta R.T. 0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56


Tgt Ion:119 Resp: 960747
Ion Ratio Lower Upper
119 100
134 25.3 12.7 38.0
91 25.0 12.7 38.1





#87
1,3-Dichlorobenzene
Concen: 105.943 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100




Tgt Ion:146 Resp: 502960
Ion Ratio Lower Upper
146 100
111 42.8 21.4 64.3
148 63.6 32.3 96.9

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#88
1,4-Dichlorobenzene
Concen: 101.341 ug/l
RT: 13.811 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56



Tgt Ion:146 Resp: 498709
Ion Ratio Lower Upper
146 100
111 40.1 21.3 63.7
148 63.1 32.4 97.0









#89
n-Butylbenzene
Concen: 123.781 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 91 Resp: 842368
Ion Ratio Lower Upper
91 100
92 52.8 25.8 77.3
134 24.3 11.7 35.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#90
Hexachloroethane
Concen: 104.405 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56


Tgt Ion:117 Resp: 188541
Ion Ratio Lower Upper
117 100
201 67.2 33.7 101.0





#91
1,2-Dichlorobenzene
Concen: 99.445 ug/l
RT: 14.100 min Scan# 2065
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 146 Resp: 470919
Ion Ratio Lower Upper
146 100
111 43.9 21.7 65.1
148 64.0 31.4 94.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 100.313 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 75 Resp: 63846
Ion Ratio Lower Upper
75 100
155 73.5 36.4 109.2
157 93.0 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 113.946 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion:180 Resp: 250787

Ion Ratio Lower Upper

180 100

182 94.9 46.8 140.3





145 32.6 16.0 48.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#94

Hexachlorobutadiene

Concen: 96.619 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN085438.D

Acq: 14 Jan 2025 14:56




Tgt Ion:225 Resp: 112899


Ion Ratio Lower Upper

225 100

223 61.8 30.7 92.1


227 63.6 30.9 92.5





#95
Naphthalene
Concen: 115.234 ug/l
RT: 15.635 min Scan# 2326
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion:128 Resp: 756816
Ion Ratio Lower Upper
128 100
127 13.1 10.6 16.0
129 10.8 8.8 13.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#96
1,2,3-Trichlorobenzene
Concen: 107.256 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:180 Resp: 238860
Ion Ratio Lower Upper
180 100
182 95.0 47.4 142.2
145 34.3 16.9 50.7

