

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085443.D
 Acq On : 14 Jan 2025 17:19
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Quant Time: Jan 15 01:46:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	166880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	302938	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	256851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	99539	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.124	85	2382	1.054	ug/l	91
3) Chloromethane	2.359	50	2801	1.145	ug/l	93
4) Vinyl Chloride	2.512	62	2732	1.111	ug/l #	81
6) Chloroethane	3.124	64	1809	1.160	ug/l	94
7) Trichlorofluoromethane	3.512	101	3860	1.082	ug/l	88
8) Diethyl Ether	3.953	74	1526	1.238	ug/l	81
9) 1,1,2-Trichlorotrifluo...	4.359	101	2156m	1.073	ug/l	
12) 1,1-Dichloroethene	4.330	96	1660	0.927	ug/l #	67
14) Allyl chloride	5.024	41	3179	1.094	ug/l #	85
15) Acrylonitrile	5.718	53	4967	5.070	ug/l #	70
16) Acetone	4.436	43	5102	5.862	ug/l	92
17) Carbon Disulfide	4.706	76	6601	1.197	ug/l #	87
18) Methyl Acetate	5.030	43	2908	1.099	ug/l #	80
19) Methyl tert-butyl Ether	5.789	73	5156	0.887	ug/l #	87
20) Methylene Chloride	5.265	84	2190	1.016	ug/l #	87
21) trans-1,2-Dichloroethene	5.794	96	2110	1.102	ug/l #	66
22) Diisopropyl ether	6.677	45	5691	0.882	ug/l #	88
23) Vinyl Acetate	6.606	43	18526m	4.090	ug/l	
24) 1,1-Dichloroethane	6.583	63	4020m	1.022	ug/l	
25) 2-Butanone	7.483	43	6448m	5.034	ug/l	
26) 2,2-Dichloropropane	7.488	77	3104	0.976	ug/l	92
27) cis-1,2-Dichloroethene	7.477	96	2185	0.969	ug/l	92
28) Bromochloromethane	7.812	49	2082	1.138	ug/l #	89
29) Tetrahydrofuran	7.835	42	3693	4.548	ug/l	98
30) Chloroform	7.965	83	4250	1.045	ug/l #	80
32) 1,1,1-Trichloroethane	8.165	97	3678	1.031	ug/l #	56
36) 1,1-Dichloropropene	8.371	75	3084	1.046	ug/l #	84
37) Ethyl Acetate	7.577	43	3056m	1.026	ug/l	
38) Carbon Tetrachloride	8.359	117	3436	1.018	ug/l	90
39) Methylcyclohexane	9.600	83	2403	0.867	ug/l #	87
40) Benzene	8.606	78	8480	0.957	ug/l	91
41) Methacrylonitrile	7.771	41	1458	0.941	ug/l #	76
42) 1,2-Dichloroethane	8.665	62	3131	0.938	ug/l	91
43) Isopropyl Acetate	8.682	43	4803	1.007	ug/l #	87
44) Trichloroethene	9.347	130	2133	1.034	ug/l	71
45) 1,2-Dichloropropane	9.618	63	2246	0.992	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
 Data File : VN085443.D
 Acq On : 14 Jan 2025 17:19
 Operator : JC\MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001

Quant Time: Jan 15 01:46:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 01:28:51 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	9.706	93	1769	1.083	ug/l	92
47) Bromodichloromethane	9.882	83	2935	0.882	ug/l	90
48) Methyl methacrylate	9.677	41	1886	0.878	ug/l	96
49) 1,4-Dioxane	9.688	88	673	18.616	ug/l #	47
51) 4-Methyl-2-Pentanone	10.441	43	11523	4.163	ug/l	93
52) Toluene	10.618	92	4180	0.814	ug/l	93
53) t-1,3-Dichloropropene	10.835	75	2522	0.802	ug/l	92
54) cis-1,3-Dichloropropene	10.312	75	2725	0.811	ug/l #	83
55) 1,1,2-Trichloroethane	11.018	97	1901	0.935	ug/l	91
56) Ethyl methacrylate	10.871	69	1923	2.471	ug/l #	88
57) 1,3-Dichloropropane	11.165	76	3113	0.881	ug/l	95
58) 2-Chloroethyl Vinyl ether	10.153	63	4929	3.822	ug/l	91
59) 2-Hexanone	11.194	43	7905	4.058	ug/l	87
60) Dibromochloromethane	11.359	129	2339	0.953	ug/l	98
61) 1,2-Dibromoethane	11.465	107	2022	1.000	ug/l	90
64) Tetrachloroethene	11.106	164	1658	0.947	ug/l #	83
65) Chlorobenzene	11.888	112	5401	0.960	ug/l	93
66) 1,1,1,2-Tetrachloroethane	11.959	131	2185	1.058	ug/l #	64
67) Ethyl Benzene	11.959	91	7347	0.802	ug/l	94
68) m/p-Xylenes	12.070	106	5059	1.494	ug/l	100
69) o-Xylene	12.394	106	2478	0.766	ug/l	97
70) Styrene	12.406	104	3812	0.712	ug/l	99
71) Bromoform	12.582	173	1208	0.818	ug/l #	88
73) Isopropylbenzene	12.688	105	5506	0.820	ug/l	98
74) N-amyl acetate	12.494	43	2716	0.899	ug/l #	87
75) 1,1,1,2-Tetrachloroethane	12.935	83	2616	1.102	ug/l #	96
76) 1,2,3-Trichloropropane	12.982	75	2198m	1.088	ug/l	
77) Bromobenzene	12.976	156	1733	0.987	ug/l	97
78) n-propylbenzene	13.029	91	6424	0.808	ug/l	93
79) 2-Chlorotoluene	13.117	91	4483	0.871	ug/l	97
80) 1,3,5-Trimethylbenzene	13.165	105	4301	0.775	ug/l	90
82) 4-Chlorotoluene	13.223	91	4268	0.833	ug/l	99
83) tert-Butylbenzene	13.435	119	3753	0.805	ug/l	96
84) 1,2,4-Trimethylbenzene	13.482	105	3758	0.679	ug/l	93
85) sec-Butylbenzene	13.612	105	4614	0.714	ug/l	99
86) p-Isopropyltoluene	13.723	119	3538	0.716	ug/l	98
87) 1,3-Dichlorobenzene	13.723	146	3038	0.940	ug/l	95
88) 1,4-Dichlorobenzene	13.806	146	3517m	1.050	ug/l	
89) n-Butylbenzene	14.053	91	3527	0.761	ug/l	93
90) Hexachloroethane	14.341	117	1355	1.102	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	3515	1.090	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	14.711	75	402	0.928	ug/l	74
93) 1,2,4-Trichlorobenzene	15.394	180	1310	0.874	ug/l #	80
94) Hexachlorobutadiene	15.500	225	822	1.033	ug/l	89
95) Naphthalene	15.641	128	4210	0.941	ug/l #	90
96) 1,2,3-Trichlorobenzene	15.835	180	1494	0.985	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085443.D
Acq On : 14 Jan 2025 17:19
Operator : JC\MD
Sample : VSTDIC001
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025


Quant Time: Jan 15 01:46:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion:168 Resp: 166880
Ion Ratio Lower Upper
168 100
99 68.6 53.6 80.4




Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#2
Dichlorodifluoromethane
Concen: 1.054 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 85 Resp: 2382
Ion Ratio Lower Upper
85 100
87 28.5 16.7 50.1




Abundance Scan 69 (2.359 min): VN085439.D\data.ms (-62)



#3
Chloromethane
Concen: 1.145 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 50 Resp: 2801
Ion Ratio Lower Upper
50 100
52 29.7 27.0 40.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#4
Vinyl Chloride
Concen: 1.111 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 62 Resp: 2732
Ion Ratio Lower Upper
62 100
64 41.4 24.8 37.2#









#6
Chloroethane
Concen: 1.160 ug/l
RT: 3.124 min Scan# 198
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion: 64 Resp: 1809
Ion Ratio Lower Upper
64 100
66 34.1 24.6 36.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#7
Trichlorofluoromethane
Concen: 1.082 ug/l
RT: 3.512 min Scan# 265
Delta R.T. 0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19


Tgt Ion:101 Resp: 3860
Ion Ratio Lower Upper
101 100
103 55.0 51.4 77.2





#8
Diethyl Ether
Concen: 1.238 ug/l
RT: 3.953 min Scan# 340
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 74 Resp: 1526
Ion Ratio Lower Upper
74 100
45 80.9 49.7 149.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#9
1,1,2-Trichlorotrifluoroethane
Concen: 1.073 ug/l m
RT: 4.359 min Scan# 409
Delta R.T. -0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19


Tgt Ion:101 Resp: 2156
Ion Ratio Lower Upper
101 100
85 35.3 37.8 56.8#
151 66.4 58.8 88.2





#12
1,1-Dichloroethene
Concen: 0.927 ug/l
RT: 4.330 min Scan# 406
Delta R.T. -0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001




Tgt Ion: 96 Resp: 1660
Ion Ratio Lower Upper
96 100
61 122.1 141.0 211.4#
98 48.1 49.0 73.6#

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#14
Allyl chloride
Concen: 1.094 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Tgt Ion: 41 Resp: 3179
Ion Ratio Lower Upper
41 100
39 73.1 64.4 96.6
76 21.4 30.5 45.7#





#15
Acrylonitrile
Concen: 5.070 ug/l
RT: 5.718 min Scan# 64
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 53 Resp: 4967
Ion Ratio Lower Upper
53 100
52 45.1 65.5 98.3#
51 33.3 29.8 44.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#16
Acetone
Concen: 5.862 ug/l
RT: 4.436 min Scan# 422
Delta R.T. 0.018 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19


Tgt Ion: 43 Resp: 5102
Ion Ratio Lower Upper
43 100
58 33.8 23.8 35.6





#17
Carbon Disulfide
Concen: 1.197 ug/l
RT: 4.706 min Scan# 468
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument : MSVOA_N
ClientSampleId : VSTDIC001




Tgt Ion: 76 Resp: 6601
Ion Ratio Lower Upper
76 100
78 13.2 6.9 10.3#

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#18
Methyl Acetate
Concen: 1.099 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19



Tgt Ion: 43 Resp: 2908
Ion Ratio Lower Upper
43 100
74 13.8 18.6 28.0#





#19

Methyl tert-butyl Ether

Concen: 0.887 ug/l

RT: 5.789 min Scan# 65

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 73 Resp: 5156

Ion Ratio Lower Upper

73 100





57 28.4 17.8 26.8#

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#20

Methylene Chloride

Concen: 1.016 ug/l

RT: 5.265 min Scan# 563

Delta R.T. -0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion: 84 Resp: 2190




Ion Ratio Lower Upper


84 100

49 115.6 104.1 156.1

51 36.9 31.0 46.4

86 48.9 51.9 77.9#





#21

trans-1,2-Dichloroethene

Concen: 1.102 ug/l

RT: 5.794 min Scan# 651

Delta R.T. 0.012 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 96 Resp: 2110

Ion Ratio Lower Upper

96 100

61 98.7 115.7 173.5#





98 41.8 49.8 74.8#

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#22

Diisopropyl ether

Concen: 0.882 ug/l

RT: 6.677 min Scan# 803

Delta R.T. 0.012 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion: 45 Resp: 5691




Ion Ratio Lower Upper


45 100

43 65.9 45.0 67.6

87 35.0 21.8 32.8#

59 11.3 10.2 15.2





#23

Vinyl Acetate

Concen: 4.090 ug/l m

RT: 6.606 min Scan# 791

Delta R.T. 0.012 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 43 Resp: 18526

Ion Ratio Lower Upper

43 100





86 9.8 8.4 12.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#24

1,1-Dichloroethane

Concen: 1.022 ug/l m

RT: 6.583 min Scan# 787

Delta R.T. 0.018 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion: 63 Resp: 4020


Ion Ratio Lower Upper

63 100

98 0.0 3.3 9.8#

100 0.0 2.0 6.0#









#25
2-Butanone
Concen: 5.034 ug/l m
RT: 7.483 min Scan# 940
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion: 43 Resp: 6448
Ion Ratio Lower Upper
43 100
72 18.0 20.2 30.4#




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#26
2,2-Dichloropropane
Concen: 0.976 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19


Tgt Ion: 77 Resp: 3104
Ion Ratio Lower Upper
77 100
97 17.7 10.7 32.1





#27
 cis-1,2-Dichloroethene
 Concen: 0.969 ug/l
 RT: 7.477 min Scan# 931
 Delta R.T. -0.006 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001



Tgt Ion: 96 Resp: 2185
 Ion Ratio Lower Upper
 96 100
 61 166.1 0.0 311.8
 98 68.7 0.0 126.0




Manual Integrations
 APPROVED


Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025



#28
 Bromochloromethane
 Concen: 1.138 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19

Tgt Ion: 49 Resp: 2082
 Ion Ratio Lower Upper
 49 100
 129 0.0 0.0 4.4
 130 78.0 55.0 82.4









#29
Tetrahydrofuran
Concen: 4.548 ug/l
RT: 7.835 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion	Ratio	Lower	Upper
42	100		
72	43.1	34.2	51.2
71	37.9	32.5	48.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#30
Chloroform
Concen: 1.045 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19


Tgt Ion	Ratio	Lower	Upper
83	100		
85	80.2	51.8	77.6





#32
 1,1,1-Trichloroethane
 Concen: 1.031 ug/l
 RT: 8.165 min Scan# 10
 Delta R.T. -0.000 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001



Tgt Ion:	97	Resp:	3678
Ion	Ratio	Lower	Upper
97	100		
99	0.0	49.8	74.6
61	52.2	41.4	62.2

Manual Integrations
 APPROVED


Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025




#33
 1,2-Dichloroethane-d4
 Concen: 0.000 ug/l
 Expected RT: 8.57 min

Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19


Tgt Ion:	65
Sig	Exp Ratio
65	100
67	50.8





#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1214
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 114 Resp: 302938
Ion Ratio Lower Upper
114 100
63 22.3 0.0 47.6
88 16.8 0.0 32.6

Manual Integrations
APPROVED


Reviewed By : John Carlone 01/15/2025
Supervised By : Mahesh Dadoda 01/15/2025




#35
Dibromofluoromethane
Concen: 0.000 ug/l
Expected RT: 8.16 min

Lab File: VN085443.D
Acq: 14 Jan 2025 17:19


Tgt Ion: 113
Sig Exp Ratio
113 100
111 103.4
192 17.9





#36
1,1-Dichloropropene
Concen: 1.046 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001




Tgt Ion: 75 Resp: 3084
Ion Ratio Lower Upper
75 100
110 30.1 16.5 49.5
77 15.9 24.4 36.6#

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#37
Ethyl Acetate
Concen: 1.026 ug/l m
RT: 7.577 min Scan# 956
Delta R.T. 0.023 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Tgt Ion: 43 Resp: 3056
Ion Ratio Lower Upper
43 100
61 4.3 10.9 16.3#
70 0.0 7.5 11.3#





#38
Carbon Tetrachloride
Concen: 1.018 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion: 117 Resp: 3436
Ion Ratio Lower Upper
117 100
119 106.6 75.4 113.2
121 31.0 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#39
Methylcyclohexane
Concen: 0.867 ug/l
RT: 9.600 min Scan# 1300
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 83 Resp: 2403
Ion Ratio Lower Upper
83 100
55 72.7 62.6 94.0
98 31.8 37.7 56.5#









#40
Benzene
Concen: 0.957 ug/l
RT: 8.606 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion: 78 Resp: 8480
Ion Ratio Lower Upper
78 100
77 28.3 19.0 28.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#41
Methacrylonitrile
Concen: 0.941 ug/l
RT: 7.771 min Scan# 989
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 41 Resp: 1458
Ion Ratio Lower Upper
41 100
39 37.6 46.0 69.0#
67 53.8 57.4 86.2#
52 18.5 25.5 38.3#









#42
1,2-Dichloroethane
Concen: 0.938 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion: 62 Resp: 3131
Ion Ratio Lower Upper
62 100
98 5.4 0.0 17.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#43
Isopropyl Acetate
Concen: 1.007 ug/l
RT: 8.682 min Scan# 1144
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19


Tgt Ion: 43 Resp: 4803
Ion Ratio Lower Upper
43 100
61 18.9 20.7 31.1#
87 7.3 9.8 14.8#





#44
 Trichloroethene
 Concen: 1.034 ug/l
 RT: 9.347 min Scan# 1257
 Delta R.T. -0.000 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19




Instrument : MSVOA_N
 ClientSampleId : VSTDIC001




Tgt Ion: 130 Resp: 2133
 Ion Ratio Lower Upper
 130 100
 95 126.1 0.0 195.8

Manual Integrations
 APPROVED



Reviewed By : John Carlone 01/15/2025
 Supervised By : Mahesh Dadoda 01/15/2025




#45
 1,2-Dichloropropane
 Concen: 0.992 ug/l
 RT: 9.618 min Scan# 1303
 Delta R.T. -0.000 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19



Tgt Ion: 63 Resp: 2246
 Ion Ratio Lower Upper
 63 100
 65 32.6 25.6 38.4





#46

Dibromomethane

Concen: 1.083 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN085443.D


Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001






Tgt Ion:	93	Resp:	1769
Ion	Ratio	Lower	Upper
93	100		
95	86.5	64.7	97.1
174	77.4	69.0	103.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#47

Bromodichloromethane

Concen: 0.882 ug/l




RT: 9.882 min Scan# 1348


Delta R.T. 0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19


Tgt Ion:	83	Resp:	2935
Ion	Ratio	Lower	Upper
83	100		
85	72.9	51.2	76.8
127	9.0	6.5	9.7





#48
Methyl methacrylate
Concen: 0.878 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion: 41 Resp: 1886
Ion Ratio Lower Upper
41 100
69 76.2 64.7 97.1
39 62.8 49.0 73.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#49
1,4-Dioxane
Concen: 18.616 ug/l
RT: 9.688 min Scan# 1315
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 88 Resp: 673
Ion Ratio Lower Upper
88 100
43 0.0 26.6 39.8#
58 31.1 59.5 89.3#





#50
 Toluene-d8
 Concen: 0.000 ug/l
 Expected RT: 10.56 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19

Tgt Ion: 98
 Sig Exp Ratio
 98 100
 100 65.3

Instrument :

MSVOA_N

ClientSampleId :



VSTDICC001

Manual Integrations

APPROVED




Reviewed By :John Carlone 01/15/2025


Supervised By :Mahesh Dadoda 01/15/2025



#51
 4-Methyl-2-Pentanone
 Concen: 4.163 ug/l
 RT: 10.441 min Scan# 1443
 Delta R.T. 0.000 min
 Lab File: VN085443.D
 Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 11523
 Ion Ratio Lower Upper
 43 100
 58 33.6 30.5 45.7





#52

Toluene

Concen: 0.814 ug/l

RT: 10.618 min Scan# 1473

Delta R.T. -0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 92 Resp: 4180

Ion Ratio Lower Upper

92 100





91 183.9 139.2 208.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#53

t-1,3-Dichloropropene

Concen: 0.802 ug/l

RT: 10.835 min Scan# 1510

Delta R.T. -0.000 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion: 75 Resp: 2522

Ion Ratio Lower Upper

75 100

77 33.5 23.5 35.3





#54

cis-1,3-Dichloropropene

Concen: 0.811 ug/l

RT: 10.312 min Scan# 14

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt	Ion: 75	Resp: 2725
Ion	Ratio	Lower Upper
75	100	
77	39.0	25.0 37.4
39	67.6	43.1 64.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#55

1,1,2-Trichloroethane

Concen: 0.935 ug/l




RT: 11.018 min Scan# 1541


Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt	Ion: 97	Resp: 1901
Ion	Ratio	Lower Upper
97	100	
83	101.5	69.0 103.6
85	60.3	44.6 66.8
99	59.9	48.6 72.8









#56
Ethyl methacrylate
Concen: 2.471 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion:	69	Resp:	1923
Ion	Ratio	Lower	Upper
69	100		
41	74.4	54.6	82.0
39	52.4	32.4	48.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#57
1,3-Dichloropropane
Concen: 0.881 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:	76	Resp:	3113
Ion	Ratio	Lower	Upper
76	100		
78	34.9	25.6	38.4









#58
2-Chloroethyl Vinyl ether
Concen: 3.822 ug/l
RT: 10.153 min Scan# 1395
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Instrument : MSVOA_N
ClientSampleId : VSTDIC001

Tgt Ion: 63 Resp: 4929
Ion Ratio Lower Upper
63 100
106 31.7 21.6 32.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#59
2-Hexanone
Concen: 4.058 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 7905
Ion Ratio Lower Upper
43 100
58 42.9 26.2 78.6





#60

Dibromochloromethane

Concen: 0.953 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion:129 Resp: 2339

Ion Ratio Lower Upper

129 100





127 78.9 38.6 115.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#61

1,2-Dibromoethane

Concen: 1.000 ug/l

RT: 11.465 min Scan# 1617

Delta R.T. -0.000 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion:107 Resp: 2022

Ion Ratio Lower Upper

107 100

109 84.8 75.9 113.9





#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 12.85 min



Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 95
Sig Exp Ratio
95 100
174 72.5
176 71.2

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001




Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19


Tgt Ion:117 Resp: 256851
Ion Ratio Lower Upper
117 100
82 59.0 48.6 72.8
119 32.2 26.6 39.8





#64
Tetrachloroethene
Concen: 0.947 ug/l
RT: 11.106 min Scan# 1555
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion:164 Resp: 1658
Ion Ratio Lower Upper
164 100
166 152.1 103.4 155.2
129 128.3 79.2 118.8#
131 96.9 77.1 115.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#65
Chlorobenzene
Concen: 0.960 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19


Tgt Ion:112 Resp: 5401
Ion Ratio Lower Upper
112 100
114 35.3 25.3 37.9





#66
1,1,1,2-Tetrachloroethane
Concen: 1.058 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion	Ratio	Lower	Upper
131	100		
133	97.7	47.4	142.3
119	0.0	33.1	99.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#67
Ethyl Benzene
Concen: 0.802 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion	Ratio	Lower	Upper
91	100		
106	32.8	23.8	35.8









#68
m/p-Xylenes
Concen: 1.494 ug/l
RT: 12.070 min Scan# 1719
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion:106 Resp: 5059
Ion Ratio Lower Upper
106 100
91 209.3 167.7 251.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#69
o-Xylene
Concen: 0.766 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:106 Resp: 2478
Ion Ratio Lower Upper
106 100
91 225.4 110.4 331.2





#70

Styrene

Concen: 0.712 ug/l

RT: 12.406 min Scan# 17

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	104	Resp:	3812
Ion Ratio	Lower	Upper	
104	100		
78	52.9	42.5	63.7
103	53.7	43.8	65.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#71

Bromoform

Concen: 0.818 ug/l




RT: 12.582 min Scan# 1807


Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion:	173	Resp:	1208
Ion Ratio	Lower	Upper	
173	100		
175	40.7	24.4	73.2
254	0.0	0.0	0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion	Ratio	Lower	Upper
152	100		
115	63.8	31.1	93.3
150	158.5	0.0	343.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#73

Isopropylbenzene

Concen: 0.820 ug/l




RT: 12.688 min Scan# 1825


Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion	Ratio	Lower	Upper
105	100		
120	26.7	12.8	38.3





#74

N-amyI acetate

Concen: 0.899 ug/l

RT: 12.494 min Scan# 1791

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 43 Resp: 2716

Ion Ratio Lower Upper

43 100

70 30.4 34.0 51.0#

55 24.2 21.4 32.2





61 17.6 19.0 28.4#

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#75

1,1,2,2-Tetrachloroethane

Concen: 1.102 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion: 83 Resp: 2616


Ion Ratio Lower Upper

83 100

131 2.4 4.8 14.4#

85 65.7 32.2 96.6





#76

1,2,3-Trichloropropane

Concen: 1.088 ug/l m

RT: 12.982 min Scan# 18

Delta R.T. -0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 75 Resp: 2198

Ion Ratio Lower Upper

75 100





77 184.2 109.7 329.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#77

Bromobenzene

Concen: 0.987 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19




Tgt Ion:156 Resp: 1733


Ion Ratio Lower Upper

156 100

77 233.6 114.1 342.4

158 94.9 48.9 146.8





#78

n-propylbenzene

Concen: 0.808 ug/l

RT: 13.029 min Scan# 18

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 91 Resp: 6424

Ion Ratio Lower Upper

91 100





120 18.2 10.9 32.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#79

2-Chlorotoluene

Concen: 0.871 ug/l

RT: 13.117 min Scan# 1898

Delta R.T. -0.006 min

Lab File: VN085443.D




Acq: 14 Jan 2025 17:19


Tgt Ion: 91 Resp: 4483

Ion Ratio Lower Upper

91 100

126 32.9 15.7 47.1





#80

1,3,5-Trimethylbenzene

Concen: 0.775 ug/l

RT: 13.165 min Scan# 19

Delta R.T. -0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC001

Tgt Ion: 105 Resp: 4301
 Ion Ratio Lower Upper
 105 100
 120 41.2 23.9 71.7

Manual Integrations
 APPROVED

Reviewed By :John Carlone 01/15/2025
 Supervised By :Mahesh Dadoda 01/15/2025



#82

4-Chlorotoluene

Concen: 0.833 ug/l




RT: 13.223 min Scan# 1916


Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion: 91 Resp: 4268
 Ion Ratio Lower Upper
 91 100
 126 31.0 15.9 47.7





#83

tert-Butylbenzene

Concen: 0.805 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC001

Tgt Ion:	119	Resp:	3753
Ion	Ratio	Lower	Upper
119	100		
91	65.9	35.1	105.3
134	24.9	12.0	36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#84

1,2,4-Trimethylbenzene

Concen: 0.679 ug/l




RT: 13.482 min Scan# 1960


Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion:	105	Resp:	3758
Ion	Ratio	Lower	Upper
105	100		
120	47.7	21.6	65.0









#85
sec-Butylbenzene
Concen: 0.714 ug/l
RT: 13.612 min Scan# 1982
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion:105 Resp: 4614
Ion Ratio Lower Upper
105 100
134 19.9 9.7 28.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#86
p-Isopropyltoluene
Concen: 0.716 ug/l
RT: 13.723 min Scan# 2001
Delta R.T. 0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19


Tgt Ion:119 Resp: 3538
Ion Ratio Lower Upper
119 100
134 25.6 12.7 38.0
91 23.2 12.7 38.1





#87
1,3-Dichlorobenzene
Concen: 0.940 ug/l
RT: 13.723 min Scan# 2015
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion:	146	Resp:	3038
Ion	Ratio	Lower	Upper
146	100		
111	43.0	21.4	64.3
148	70.9	32.3	96.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#88
1,4-Dichlorobenzene
Concen: 1.050 ug/l m
RT: 13.806 min Scan# 2015
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:	146	Resp:	3517
Ion	Ratio	Lower	Upper
146	100		
111	37.2	21.3	63.7
148	61.2	32.4	97.0









#89
n-Butylbenzene
Concen: 0.761 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion:	91	Resp:	3527
Ion Ratio	Lower	Upper	
91	100		
92	57.2	25.8	77.3
134	25.5	11.7	35.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#90
Hexachloroethane
Concen: 1.102 ug/l
RT: 14.341 min Scan# 2106
Delta R.T. 0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:	117	Resp:	1355
Ion Ratio	Lower	Upper	
117	100		
201	66.5	33.7	101.0





#91

1,2-Dichlorobenzene

Concen: 1.090 ug/l

RT: 14.106 min Scan# 2065

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt	Ion:146	Resp:	3515
Ion	Ratio	Lower	Upper
146	100		
111	43.4	21.7	65.1
148	57.8	31.4	94.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 0.928 ug/l




RT: 14.711 min Scan# 2169


Delta R.T. -0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt	Ion: 75	Resp:	402
Ion	Ratio	Lower	Upper
75	100		
155	48.0	36.4	109.2
157	112.7	45.4	136.1





#93

1,2,4-Trichlorobenzene

Concen: 0.874 ug/l

RT: 15.394 min Scan# 2284

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt	Ion:180	Resp:	1310
Ion	Ratio	Lower	Upper
180	100		
182	106.3	46.8	140.3
145	9.7	16.0	48.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#94

Hexachlorobutadiene

Concen: 1.033 ug/l




RT: 15.500 min Scan# 2303


Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt	Ion:225	Resp:	822
Ion	Ratio	Lower	Upper
225	100		
223	50.7	30.7	92.1
227	55.7	30.9	92.5





#95

Naphthalene

Concen: 0.941 ug/l

RT: 15.641 min Scan# 2326

Delta R.T. 0.006 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	128	Resp:	4210
Ion	Ratio	Lower	Upper
128	100		
127	11.1	10.6	16.0
129	5.3	8.8	13.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/15/2025

Supervised By :Mahesh Dadoda 01/15/2025



#96

1,2,3-Trichlorobenzene

Concen: 0.985 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN085443.D

Acq: 14 Jan 2025 17:19

Tgt Ion:	180	Resp:	1494
Ion	Ratio	Lower	Upper
180	100		
182	90.2	47.4	142.2
145	11.8	16.9	50.7

