

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011519\
 Data File : VN053410.D
 Acq On : 15 Jan 2019 16:14
 Operator : JC/SP
 Sample : K1139-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 S-9

Quant Time: Jan 16 04:13:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 15 11:52:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	980794	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1507906	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1370282	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	579494	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	466093	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
35) Dibromofluoromethane	7.59	113	420379	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
50) Toluene-d8	10.09	98	1861751	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
62) 4-Bromofluorobenzene	12.40	95	619197	48.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.04%	
Target Compounds						
16) Acetone	3.82	43	15208	6.686	ug/l	99
20) Methylene Chloride	4.56	84	599058	57.691	ug/l	99
43) Isopropyl Acetate	8.17	43	51059	3.442	ug/l #	83
95) Naphthalene	15.13	128	5450	4.297	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011519\
Data File : VN053410.D
Acq On : 15 Jan 2019 16:14
Operator : JC/SP
Sample : K1139-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
S-9

Quant Time: Jan 16 04:13:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 15 11:52:35 2019
Response via : Initial Calibration

