

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011521\
 Data File : VN065518.D
 Acq On : 15 Jan 2021 12:57
 Operator : JC/MD
 Sample : VN0115MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0115MBL01

Quant Time: Jan 15 23:39:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	189033	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	305767	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	321126	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	131204	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.990	65	97239	45.57	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.14%
35) Dibromofluoromethane	7.553	113	94543	51.25	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.50%
50) Toluene-d8	10.058	98	377765	54.64	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.28%
62) 4-Bromofluorobenzene	12.376	95	138842	50.70	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.40%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011521\
Data File : VN065518.D
Acq On : 15 Jan 2021 12:57
Operator : JC/MD
Sample : VN0115MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0115MBL01

Quant Time: Jan 15 23:39:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

