

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011521\
 Data File : VN065523.D
 Acq On : 15 Jan 2021 15:33
 Operator : JC/MD
 Sample : M1142-02 100X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30936

Quant Time: Jan 15 23:41:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	204109	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	333226	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	361671	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	160608	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	108320	47.01	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.02%		
35) Dibromofluoromethane	7.553	113	103431	51.45	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	102.90%		
50) Toluene-d8	10.058	98	414115	54.96	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	109.92%		
62) 4-Bromofluorobenzene	12.376	95	161910	53.89	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	107.78%		
Target Compounds						
16) Acetone	3.791	43	1440039	1545.02	ug/l	90
52) Toluene	10.122	92	151062	26.59	ug/l	100
95) Naphthalene	15.106	128	3262	3.74	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011521\
Data File : VN065523.D
Acq On : 15 Jan 2021 15:33
Operator : JC/MD
Sample : M1142-02 100X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30936

Quant Time: Jan 15 23:41:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

