

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011521\
 Data File : VN065529.D
 Acq On : 15 Jan 2021 18:02
 Operator : JC/MD
 Sample : M1124-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-223

Quant Time: Jan 15 23:43:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	202508	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	330159	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	369519	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	158277	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	108146	47.30	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 94.60%			
35) Dibromofluoromethane	7.553	113	102301	51.36	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.72%			
50) Toluene-d8	10.061	98	418826	56.10	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 112.20%			
62) 4-Bromofluorobenzene	12.376	95	162960	54.66	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 109.32%			
Target Compounds						
20) Methylene Chloride	4.508	84	9495	4.26	ug/l	93
43) Isopropyl Acetate	8.135	43	14005	2.73	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011521\
Data File : VN065529.D
Acq On : 15 Jan 2021 18:02
Operator : JC/MD
Sample : M1124-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-223

Quant Time: Jan 15 23:43:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

