

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011521\
 Data File : VN065531.D
 Acq On : 15 Jan 2021 18:51
 Operator : JC/MD
 Sample : M1141-21
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T872-T871

Quant Time: Jan 15 23:44:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	191776	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	317235	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	355540	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	160590	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	102614	47.40	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 94.80%			
35) Dibromofluoromethane	7.553	113	97743	51.07	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.14%			
50) Toluene-d8	10.061	98	403159	56.20	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 112.40%			
62) 4-Bromofluorobenzene	12.376	95	158905	55.40	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 110.80%			
Target Compounds						
20) Methylene Chloride	4.512	84	10663	5.05	ug/l	97
43) Isopropyl Acetate	8.135	43	15328	3.11	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011521\
Data File : VN065531.D
Acq On : 15 Jan 2021 18:51
Operator : JC/MD
Sample : M1141-21
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T872-T871

Quant Time: Jan 15 23:44:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

