

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011521\
 Data File : VN065533.D
 Acq On : 15 Jan 2021 19:41
 Operator : JC/MD
 Sample : M1146-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 520-Z

Quant Time: Jan 15 23:44:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	201015	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	336172	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	368996	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	157951	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	106402	46.89	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.78%
35) Dibromofluoromethane	7.553	113	102171	50.38	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.76%
50) Toluene-d8	10.061	98	422275	55.55	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	111.10%
62) 4-Bromofluorobenzene	12.376	95	163643	53.98	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.96%
Target Compounds						
20) Methylene Chloride	4.508	84	12683	5.73	ug/l	98
43) Isopropyl Acetate	8.132	43	14751	2.82	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011521\
Data File : VN065533.D
Acq On : 15 Jan 2021 19:41
Operator : JC/MD
Sample : M1146-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
520-Z

Quant Time: Jan 15 23:44:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

