

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011521\
 Data File : VN065534.D
 Acq On : 15 Jan 2021 20:05
 Operator : JC/MD
 Sample : M1090-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BORE-3-1FT-7FT

Quant Time: Jan 15 23:45:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	198047	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	330078	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	365138	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	154698	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	105750	47.30	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery =	94.60%		
35) Dibromofluoromethane	7.556	113	99079	49.75	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery =	99.50%		
50) Toluene-d8	10.061	98	415784	55.71	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery =	111.42%		
62) 4-Bromofluorobenzene	12.376	95	159449	53.61	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery =	107.22%		
Target Compounds						
						Qvalue
20) Methylene Chloride	4.508	84	17158	7.87	ug/l	99
43) Isopropyl Acetate	8.135	43	14873	2.90	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011521\
 Data File : VN065534.D
 Acq On : 15 Jan 2021 20:05
 Operator : JC/MD
 Sample : M1090-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BORE-3-1FT-7FT

Quant Time: Jan 15 23:45:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

