

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011524\
 Data File : VN080717.D
 Acq On : 15 Jan 2024 15:55
 Operator : JC\MD
 Sample : P1144-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SCMW-20S

Quant Time: Jan 16 00:42:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	148522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	340209	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	343712	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	143990	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	115567	44.309	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.620%
35) Dibromofluoromethane	8.165	113	96028	49.679	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.360%
50) Toluene-d8	10.565	98	393787	53.427	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.860%
62) 4-Bromofluorobenzene	12.847	95	155146	54.891	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.780%
Target Compounds						
					Qvalue	
3) Chloromethane	2.371	50	2527	0.272	ug/l	89
27) cis-1,2-Dichloroethene	7.483	96	7170	2.892	ug/l	99
44) Trichloroethene	9.347	130	9148	3.630	ug/l	98
64) Tetrachloroethene	11.106	164	32631	13.824	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011524\
Data File : VN080717.D
Acq On : 15 Jan 2024 15:55
Operator : JC\MD
Sample : P1144-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SCMW-20S

Quant Time: Jan 16 00:42:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

