

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011524\
 Data File : VN080720.D
 Acq On : 15 Jan 2024 17:06
 Operator : JC\MD
 Sample : P1144-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SCMW-29D

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/16/2024
 Supervised By : Mahesh Dadoda 01/16/2024

Quant Time: Jan 16 00:42:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	148156	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	333039	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	342148	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	140733	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	110822	42.595	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	85.180%		
35) Dibromofluoromethane	8.171	113	93250	49.280	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.560%		
50) Toluene-d8	10.565	98	378086	52.401	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	104.800%		
62) 4-Bromofluorobenzene	12.847	95	148055	53.510	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	107.020%		
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	4.336	96	1598m	0.722	ug/l	
16) Acetone	4.430	43	3749	4.389	ug/l	95
27) cis-1,2-Dichloroethene	7.465	96	2030	0.821	ug/l	# 8
44) Trichloroethene	9.347	130	16632	6.742	ug/l	91
64) Tetrachloroethene	11.106	164	5541	2.358	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011524\
Data File : VN080720.D
Acq On : 15 Jan 2024 17:06
Operator : JC\MD
Sample : P1144-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SCMW-29D

Quant Time: Jan 16 00:42:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/16/2024
Supervised By :Mahesh Dadoda 01/16/2024

