

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
 Data File : VN085449.D
 Acq On : 15 Jan 2025 10:05
 Operator : JC\MD
 Sample : VN0115WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0115WBL01

Quant Time: Jan 16 00:53:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	195111	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	360723	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	319974	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	130534	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	173401	55.057	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 110.120%		
35) Dibromofluoromethane	8.165	113	129916	51.914	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.820%		
50) Toluene-d8	10.565	98	445519	50.107	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.220%		
62) 4-Bromofluorobenzene	12.847	95	149204	49.055	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 98.120%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
Data File : VN085449.D
Acq On : 15 Jan 2025 10:05
Operator : JC\MD
Sample : VN0115WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0115WBL01

Quant Time: Jan 16 00:53:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

