

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
 Data File : VN085452.D
 Acq On : 15 Jan 2025 11:26
 Operator : JC\MD
 Sample : Q1062-04 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-190A-GW-898-900

Quant Time: Jan 16 00:55:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	175474	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	316452	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	274227	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	107831	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	149471	52.770	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.540%
35) Dibromofluoromethane	8.165	113	114177	52.007	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.020%
50) Toluene-d8	10.565	98	384211	49.257	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.520%
62) 4-Bromofluorobenzene	12.847	95	123439	46.262	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
Data File : VN085452.D
Acq On : 15 Jan 2025 11:26
Operator : JC\MD
Sample : Q1062-04 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-190A-GW-898-900

Quant Time: Jan 16 00:55:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

