

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
 Data File : VN085453.D
 Acq On : 15 Jan 2025 11:50
 Operator : JC\MD
 Sample : Q1062-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-190A-TB-20250107

Quant Time: Jan 16 00:55:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	196038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	355455	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	304386	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	119672	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	168714	53.315	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	106.640%	
35) Dibromofluoromethane	8.159	113	123337	50.015	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.040%	
50) Toluene-d8	10.565	98	428715	48.931	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	97.860%	
62) 4-Bromofluorobenzene	12.847	95	137686	45.939	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	91.880%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
Data File : VN085453.D
Acq On : 15 Jan 2025 11:50
Operator : JC\MD
Sample : Q1062-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-190A-TB-20250107

Quant Time: Jan 16 00:55:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

