

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
 Data File : VN085454.D
 Acq On : 15 Jan 2025 12:14
 Operator : JC\MD
 Sample : Q1062-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-190A-EB-20250107

Quant Time: Jan 16 00:56:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	185335	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	337105	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	298295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	117680	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	160466	53.637	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.280%			
35) Dibromofluoromethane	8.159	113	118687	50.750	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.500%			
50) Toluene-d8	10.565	98	415115	49.958	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.920%			
62) 4-Bromofluorobenzene	12.847	95	132424	46.589	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 93.180%			
Target Compounds						
29) Tetrahydrofuran	7.835	42	5881	6.521	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
Data File : VN085454.D
Acq On : 15 Jan 2025 12:14
Operator : JC\MD
Sample : Q1062-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-190A-EB-20250107

Quant Time: Jan 16 00:56:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

