

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
 Data File : VN085456.D
 Acq On : 15 Jan 2025 13:01
 Operator : JC\MD
 Sample : PB166029TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166029TB

Quant Time: Jan 16 00:57:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.223	168	181616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	348273	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	321375	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	128492	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	159448	54.388	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.780%		
35) Dibromofluoromethane	8.165	113	120640	49.930	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.860%		
50) Toluene-d8	10.564	98	444413	51.769	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.540%		
62) 4-Bromofluorobenzene	12.847	95	141316	48.123	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 96.240%		
Target Compounds						
						Qvalue
16) Acetone	4.424	43	100674	106.281	ug/l	99
25) 2-Butanone	7.482	43	9452	6.781	ug/l	93
43) Isopropyl Acetate	8.682	43	179066	32.650	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
Data File : VN085456.D
Acq On : 15 Jan 2025 13:01
Operator : JC\MD
Sample : PB166029TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166029TB

Quant Time: Jan 16 00:57:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

