

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
 Data File : VN085458.D
 Acq On : 15 Jan 2025 13:49
 Operator : JC\MD
 Sample : PB166029ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166029ZHE#02

Quant Time: Jan 16 00:58:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	185364	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	356019	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	327610	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	123822	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	163676	54.702	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	109.400%	
35) Dibromofluoromethane	8.165	113	120104	48.627	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.260%	
50) Toluene-d8	10.565	98	449038	51.170	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.340%	
62) 4-Bromofluorobenzene	12.847	95	137929	45.948	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	91.900%	
Target Compounds						
16) Acetone	4.424	43	45225	46.778	ug/l	92
20) Methylene Chloride	5.277	84	3719	1.554	ug/l	95
30) Chloroform	7.965	83	66116	14.642	ug/l	98
43) Isopropyl Acetate	8.683	43	197284	35.189	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
Data File : VN085458.D
Acq On : 15 Jan 2025 13:49
Operator : JC\MD
Sample : PB166029ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166029ZHE#02

Quant Time: Jan 16 00:58:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

