

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
 Data File : VN085459.D
 Acq On : 15 Jan 2025 14:13
 Operator : JC\MD
 Sample : PB166029ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166029ZHE#03

Quant Time: Jan 16 00:58:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	180587	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	355738	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	324503	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	124208	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	160959	55.217	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 110.440%			
35) Dibromofluoromethane	8.165	113	120299	48.745	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.480%			
50) Toluene-d8	10.565	98	447097	50.989	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.980%			
62) 4-Bromofluorobenzene	12.847	95	138891	46.305	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 92.600%			
Target Compounds						
16) Acetone	4.430	43	44214	46.942	ug/l	98
20) Methylene Chloride	5.271	84	3933	1.687	ug/l #	88
30) Chloroform	7.959	83	60373	13.724	ug/l	99
43) Isopropyl Acetate	8.682	43	199530	35.618	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
Data File : VN085459.D
Acq On : 15 Jan 2025 14:13
Operator : JC\MD
Sample : PB166029ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166029ZHE#03

Quant Time: Jan 16 00:58:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

