

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
 Data File : VN085447.D
 Acq On : 15 Jan 2025 09:03
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 16 00:51:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	109	0.00
2 T	Dichlorodifluoromethane	0.677	0.657	3.0	114	0.00
3 P	Chloromethane	0.733	0.674	8.0	109	0.00
4 C	Vinyl Chloride	0.737	0.696	5.6#	111	0.00
5 T	Bromomethane	0.445	0.401	9.9	105	0.00
6 T	Chloroethane	0.467	0.426	8.8	110	0.00
7 T	Trichlorofluoromethane	1.069	1.042	2.5	114	0.00
8 T	Diethyl Ether	0.369	0.338	8.4	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.602	0.600	0.3	121	0.00
10 T	Methyl Iodide	0.690	0.683	1.0	104	0.00
11 T	Tert butyl alcohol	0.092	0.069	25.0	81	0.00
12 CM	1,1-Dichloroethene	0.537	0.531	1.1#	109	0.00
13 T	Acrolein	0.126	0.120	4.8	94	0.00
14 T	Allyl chloride	0.871	0.831	4.6	108	0.00
15 T	Acrylonitrile	0.294	0.260	11.6	96	0.00
16 T	Acetone	0.261	0.230	11.9	100	0.00
17 T	Carbon Disulfide	1.652	1.522	7.9	113	0.00
18 T	Methyl Acetate	0.793	0.700	11.7	97	0.00
19 T	Methyl tert-butyl Ether	1.742	1.757	-0.9	103	0.00
20 T	Methylene Chloride	0.646	0.605	6.3	105	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.562	1.9	111	0.00
22 T	Diisopropyl ether	1.933	1.966	-1.7	106	0.00
23 T	Vinyl Acetate	1.353	1.403	-3.7	103	0.00
24 P	1,1-Dichloroethane	1.178	1.146	2.7	107	0.00
25 T	2-Butanone	0.384	0.337	12.2	95	0.00
26 T	2,2-Dichloropropane	0.953	1.028	-7.9	120	0.00
27 T	cis-1,2-Dichloroethene	0.675	0.674	0.1	108	0.00
28 T	Bromochloromethane	0.548	0.540	1.5	109	0.00
29 T	Tetrahydrofuran	0.243	0.225	7.4	94	0.00
30 C	Chloroform	1.218	1.163	4.5#	108	0.00
31 T	Cyclohexane	1.024	0.942	8.0	117	0.00
32 T	1,1,1-Trichloroethane	1.068	1.035	3.1	112	0.00
33 S	1,2-Dichloroethane-d4	0.807	0.819	-1.5	108	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.347	0.369	-6.3	109	0.00
36 T	1,1-Dichloropropene	0.487	0.503	-3.3	115	0.00
37 T	Ethyl Acetate	0.491	0.440	10.4	95	0.00
38 T	Carbon Tetrachloride	0.557	0.563	-1.1	112	0.00
39 T	Methylcyclohexane	0.457	0.528	-15.5	120	0.00
40 TM	Benzene	1.463	1.504	-2.8	110	0.00
41 T	Methacrylonitrile	0.256	0.253	1.2	99	0.00
42 TM	1,2-Dichloroethane	0.551	0.550	0.2	106	0.00
43 T	Isopropyl Acetate	0.787	0.743	5.6	96	0.00
44 TM	Trichloroethene	0.341	0.345	-1.2	113	0.00
45 C	1,2-Dichloropropane	0.374	0.377	-0.8#	107	0.00
46 T	Dibromomethane	0.270	0.264	2.2	104	0.00
47 T	Bromodichloromethane	0.549	0.568	-3.5	107	0.00
48 T	Methyl methacrylate	0.354	0.361	-2.0	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
 Data File : VN085447.D
 Acq On : 15 Jan 2025 09:03
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 16 00:51:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.005	16.7	87	0.00
50 S	Toluene-d8	1.232	1.356	-10.1	113	0.00
51 T	4-Methyl-2-Pentanone	0.457	0.440	3.7	94	0.00
52 CM	Toluene	0.848	0.916	-8.0#	111	0.00
53 T	t-1,3-Dichloropropene	0.519	0.555	-6.9	106	0.00
54 T	cis-1,3-Dichloropropene	0.554	0.594	-7.2	107	0.00
55 T	1,1,2-Trichloroethane	0.335	0.337	-0.6	105	0.00
56 T	Ethyl methacrylate	0.469	0.515	-9.8	100	0.00
57 T	1,3-Dichloropropane	0.583	0.591	-1.4	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.232	-8.9	102	0.00
59 T	2-Hexanone	0.321	0.312	2.8	92	0.00
60 T	Dibromochloromethane	0.405	0.408	-0.7	104	0.00
61 T	1,2-Dibromoethane	0.334	0.327	2.1	103	0.00
62 S	4-Bromofluorobenzene	0.422	0.469	-11.1	110	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.341	0.355	-4.1	117	0.00
65 PM	Chlorobenzene	1.095	1.101	-0.5	109	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.398	1.0	108	0.00
67 C	Ethyl Benzene	1.784	1.967	-10.3#	112	0.00
68 T	m/p-Xylenes	0.659	0.755	-14.6	114	0.00
69 T	o-Xylene	0.630	0.699	-11.0	109	0.00
70 T	Styrene	1.043	1.208	-15.8	110	0.00
71 P	Bromoform	0.288	0.302	-4.9	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
73 T	Isopropylbenzene	3.375	3.662	-8.5	115	0.00
74 T	N-amyl acetate	1.517	1.451	4.4	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.192	1.047	12.2	99	0.00
76 T	1,2,3-Trichloropropane	1.016	0.948	6.7	115	0.00
77 T	Bromobenzene	0.882	0.866	1.8	109	0.00
78 T	n-propylbenzene	3.995	4.446	-11.3	116	0.00
79 T	2-Chlorotoluene	2.586	2.672	-3.3	112	0.00
80 T	1,3,5-Trimethylbenzene	2.787	3.109	-11.6	114	0.00
81 T	trans-1,4-Dichloro-2-butene	0.376	0.367	2.4	102	0.00
82 T	4-Chlorotoluene	2.574	2.686	-4.4	112	0.00
83 T	tert-Butylbenzene	2.341	2.560	-9.4	113	0.00
84 T	1,2,4-Trimethylbenzene	2.778	3.150	-13.4	113	0.00
85 T	sec-Butylbenzene	3.244	3.666	-13.0	116	0.00
86 T	p-Isopropyltoluene	2.631	3.110	-18.2	117	0.00
87 T	1,3-Dichlorobenzene	1.624	1.628	-0.2	112	0.00
88 T	1,4-Dichlorobenzene	1.683	1.618	3.9	112	0.00
89 T	n-Butylbenzene	2.327	2.684	-15.3	117	0.00
90 T	Hexachloroethane	0.618	0.601	2.8	113	0.00
91 T	1,2-Dichlorobenzene	1.620	1.533	5.4	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.218	0.187	14.2	91	0.00
93 T	1,2,4-Trichlorobenzene	0.753	0.762	-1.2	105	0.00
94 T	Hexachlorobutadiene	0.400	0.379	5.3	112	0.00
95 T	Naphthalene	2.246	2.067	8.0	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011525\
Data File : VN085447.D
Acq On : 15 Jan 2025 09:03
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 16 00:51:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.762	0.720	5.5	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6