

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011619\
 Data File : VN053428.D
 Acq On : 16 Jan 2019 16:35
 Operator : JC/SP
 Sample : PB116405ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116405ZHE#11

Quant Time: Jan 17 04:30:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	951615	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1486027	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1420993	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	641737	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	495347	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
35) Dibromofluoromethane	7.59	113	431838	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
50) Toluene-d8	10.09	98	1873697	52.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.48%	
62) 4-Bromofluorobenzene	12.40	95	653376	51.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.90%	
Target Compounds						
16) Acetone	3.82	43	13527	6.130	ug/l	93
20) Methylene Chloride	4.56	84	23528	2.335	ug/l	96
43) Isopropyl Acetate	8.17	43	63887	4.563	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011619\
Data File : VN053428.D
Acq On : 16 Jan 2019 16:35
Operator : JC/SP
Sample : PB116405ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116405ZHE#11

Quant Time: Jan 17 04:30:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

