

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN011620\
 Data File : VN059729.D
 Acq On : 16 Jan 2020 14:08
 Operator : JC/MD
 Sample : L1136-10ME
 Misc : 5.27g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP022SB458_200113_23-25ME

Quant Time: Jan 17 07:44:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 14 08:58:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	180538	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	266836	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	233067	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	108742	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	94275	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
35) Dibromofluoromethane	7.56	113	80171	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
50) Toluene-d8	10.08	98	304984	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
62) 4-Bromofluorobenzene	12.40	95	106876	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
Target Compounds						
4) Vinyl Chloride	2.17	62	14885	6.294	ug/l	93
27) cis-1,2-Dichloroethene	6.80	96	39313	19.467	ug/l	77
44) Trichloroethene	8.82	130	4959	2.621	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011620\
Data File : VN059729.D
Acq On : 16 Jan 2020 14:08
Operator : JC/MD
Sample : L1136-10ME
Misc : 5.27µ/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP022SB458_200113_23-25ME

Quant Time: Jan 17 07:44:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 14 08:58:21 2020
Response via : Initial Calibration

