

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011620\
 Data File : VN059742.D
 Acq On : 16 Jan 2020 19:12
 Operator : JC/MD
 Sample : L1136-04ME
 Misc : 7.61g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP022SB457_200113_23-25ME

Quant Time: Jan 17 07:48:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 14 08:58:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	176681	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	265173	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	226201	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	101452	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	91300	46.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.06%	
35) Dibromofluoromethane	7.56	113	80924	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
50) Toluene-d8	10.08	98	302253	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
62) 4-Bromofluorobenzene	12.40	95	103254	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
Target Compounds						
4) Vinyl Chloride	2.17	62	51398	22.208	ug/l	97
27) cis-1,2-Dichloroethene	6.80	96	98399	49.790	ug/l	77
44) Trichloroethene	8.81	130	6420	3.415	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011620\
Data File : VN059742.D
Acq On : 16 Jan 2020 19:12
Operator : JC/MD
Sample : L1136-04ME
Misc : 7.61µ/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP022SB457_200113_23-25ME

Quant Time: Jan 17 07:48:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 14 08:58:21 2020
Response via : Initial Calibration

