


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN011620\  
 Data File : VN059720.D  
 Acq On : 16 Jan 2020 9:39  
 Operator : JC/MD  
 Sample : BFB  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N011320W.M  
 Title : SW846 8260  
 Last Update : Tue Jan 14 08:58:21 2020



AutoFind: Scans 3303, 3304, 3305; Background Corrected with Scan 3291

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 15           | 40           | 17.8      | 3481    | PASS             |
| 75          | 95           | 30           | 60           | 48.8      | 9523    | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 19517   | PASS             |
| 96          | 95           | 5            | 9            | 7.5       | 1470    | PASS             |
| 173         | 174          | 0.00         | 2            | 0.4       | 69      | PASS             |
| 174         | 95           | 50           | 100          | 87.1      | 17005   | PASS             |
| 175         | 174          | 5            | 9            | 7.4       | 1258    | PASS             |
| 176         | 174          | 95           | 101          | 98.4      | 16736   | PASS             |
| 177         | 176          | 5            | 9            | 6.5       | 1084    | PASS             |