

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011625\
 Data File : VN085474.D
 Acq On : 16 Jan 2025 16:57
 Operator : JC\MD
 Sample : Q1060-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Jan 17 01:07:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	172349	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	315259	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	279617	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	110216	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	154661	55.592	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.180%
35) Dibromofluoromethane	8.165	113	112895	51.618	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.240%
50) Toluene-d8	10.565	98	383264	49.321	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.640%
62) 4-Bromofluorobenzene	12.847	95	125750	47.307	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.620%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011625\
Data File : VN085474.D
Acq On : 16 Jan 2025 16:57
Operator : JC\MD
Sample : Q1060-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Jan 17 01:07:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

