

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011625\
 Data File : VN085475.D
 Acq On : 16 Jan 2025 17:21
 Operator : JC\MD
 Sample : Q1109-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TAP-TB-01-011525

Quant Time: Jan 17 01:07:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	167864	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	311959	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	269876	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	107148	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	154079	56.863	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	113.720%	
35) Dibromofluoromethane	8.159	113	114727	53.011	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	106.020%	
50) Toluene-d8	10.565	98	382098	49.691	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	99.380%	
62) 4-Bromofluorobenzene	12.847	95	123340	46.891	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	93.780%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011625\
Data File : VN085475.D
Acq On : 16 Jan 2025 17:21
Operator : JC\MD
Sample : Q1109-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TAP-TB-01-011525

Quant Time: Jan 17 01:07:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

