

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011719\
 Data File : VN053438.D
 Acq On : 17 Jan 2019 16:53
 Operator : JC/SP
 Sample : PB116439TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116439TB

Quant Time: Jan 18 00:34:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	904256	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1406739	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1347154	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	575008	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	454553	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
35) Dibromofluoromethane	7.59	113	409700	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
50) Toluene-d8	10.09	98	1768235	52.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.16%	
62) 4-Bromofluorobenzene	12.40	95	609615	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
Target Compounds						
16) Acetone	3.82	43	11568	5.517	ug/l	99
20) Methylene Chloride	4.56	84	16290	1.702	ug/l	95
43) Isopropyl Acetate	8.17	43	37582	2.836	ug/l #	86
95) Naphthalene	15.14	128	12928	4.796	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011719\
Data File : VN053438.D
Acq On : 17 Jan 2019 16:53
Operator : JC/SP
Sample : PB116439TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116439TB

Quant Time: Jan 18 00:34:38 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

