

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011719\
 Data File : VN053444.D
 Acq On : 17 Jan 2019 19:33
 Operator : JC/SP
 Sample : PB116439ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116439ZHE#06

Quant Time: Jan 18 01:01:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	913710	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1432188	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1377175	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	614203	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	466938	51.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.40%	
35) Dibromofluoromethane	7.59	113	409813	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
50) Toluene-d8	10.09	98	1795684	51.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.90%	
62) 4-Bromofluorobenzene	12.40	95	633076	52.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.46%	
Target Compounds						
16) Acetone	3.82	43	12722	6.004	ug/l	91
20) Methylene Chloride	4.56	84	19449	2.010	ug/l	98
25) 2-Butanone	6.85	43	6945	2.127	ug/l	99
43) Isopropyl Acetate	8.17	43	53996	4.002	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011719\
Data File : VN053444.D
Acq On : 17 Jan 2019 19:33
Operator : JC/SP
Sample : PB116439ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116439ZHE#06

Quant Time: Jan 18 01:01:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

