

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011719\
 Data File : VN053450.D
 Acq On : 17 Jan 2019 22:13
 Operator : JC/SP
 Sample : PB116439ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116439ZHE#12

Quant Time: Jan 18 01:18:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	909905	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1430034	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1370613	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	610063	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	471984	52.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.96%	
35) Dibromofluoromethane	7.59	113	415012	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
50) Toluene-d8	10.09	98	1785786	51.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.48%	
62) 4-Bromofluorobenzene	12.40	95	633504	52.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.68%	
Target Compounds						
16) Acetone	3.82	43	12696	6.017	ug/l	98
20) Methylene Chloride	4.56	84	17222	1.788	ug/l	97
43) Isopropyl Acetate	8.17	43	42302	3.140	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011719\
Data File : VN053450.D
Acq On : 17 Jan 2019 22:13
Operator : JC/SP
Sample : PB116439ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116439ZHE#12

Quant Time: Jan 18 01:18:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

